

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
 Data File : VX020842.D
 Acq On : 03 Feb 2021 15:53
 Operator : JC/SP
 Sample : M1289-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-6-PD

Quant Time: Feb 04 02:23:48 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	273798	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	463368	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	449134	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	195830	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	186681	49.99	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.98%
35) Dibromofluoromethane	5.465	113	162338	51.15	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.30%
50) Toluene-d8	8.696	98	580635	49.90	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.80%
62) 4-Bromofluorobenzene	11.122	95	206805	45.76	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.52%
Target Compounds						
27) cis-1,2-Dichloroethene	4.568	96	3556	0.99	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
Data File : VX020842.D
Acq On : 03 Feb 2021 15:53
Operator : JC/SP
Sample : M1289-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-6-PD

Quant Time: Feb 04 02:23:48 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

