

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
 Data File : VX020843.D
 Acq On : 03 Feb 2021 16:17
 Operator : JC/SP
 Sample : M1289-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-7-PS

Quant Time: Feb 04 02:24:09 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	270666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	468842	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	453850	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	195566	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	188384	51.03	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.06%			
35) Dibromofluoromethane	5.464	113	162217	50.51	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.02%			
50) Toluene-d8	8.696	98	583839	49.59	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 99.18%			
62) 4-Bromofluorobenzene	11.122	95	207254	45.33	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 90.66%			
Target Compounds						
27) cis-1,2-Dichloroethene	4.568	96	9933	2.79	ug/l	87
64) Tetrachloroethene	9.317	164	794	0.23	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
Data File : VX020843.D
Acq On : 03 Feb 2021 16:17
Operator : JC/SP
Sample : M1289-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-7-PS

Quant Time: Feb 04 02:24:09 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

