

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
 Data File : VX020845.D
 Acq On : 03 Feb 2021 17:04
 Operator : JC/SP
 Sample : M1289-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-8-PS

Quant Time: Feb 04 02:24:51 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	262963	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	452025	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	430440	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	178443	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	183486	51.16	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.32%	
35) Dibromofluoromethane	5.464	113	160363	51.79	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.58%	
50) Toluene-d8	8.696	98	559350	49.28	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.56%	
62) 4-Bromofluorobenzene	11.122	95	188153	42.68	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 85.36%	
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.568	96	7568	2.19	ug/l	89
30) Chloroform	5.178	83	2582	0.45	ug/l #	71
44) Trichloroethene	7.190	130	1413	0.40	ug/l	96
64) Tetrachloroethene	9.317	164	770	0.23	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
Data File : VX020845.D
Acq On : 03 Feb 2021 17:04
Operator : JC/SP
Sample : M1289-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-8-PS

Quant Time: Feb 04 02:24:51 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

