

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
 Data File : VX020847.D
 Acq On : 03 Feb 2021 17:50
 Operator : JC/SP
 Sample : M1289-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-9-PS

Quant Time: Feb 04 02:25:37 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	270951	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	454498	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	451255	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	195664	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	185762	50.26	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.52%	
35) Dibromofluoromethane	5.465	113	160470	51.55	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.10%	
50) Toluene-d8	8.696	98	567500	49.72	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.44%	
62) 4-Bromofluorobenzene	11.122	95	201197	45.39	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 90.78%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.568	96	5457	1.53	ug/l	94
44) Trichloroethene	7.190	130	1632	0.46	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
Data File : VX020847.D
Acq On : 03 Feb 2021 17:50
Operator : JC/SP
Sample : M1289-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-9-PS

Quant Time: Feb 04 02:25:37 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

