

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
 Data File : VX020851.D
 Acq On : 03 Feb 2021 19:24
 Operator : JC/SP
 Sample : M1289-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-FD-02

Quant Time: Feb 04 02:27:00 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	275286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	476446	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	461218	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	204394	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	193612	51.56	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	103.12%	
35) Dibromofluoromethane	5.464	113	167107	51.20	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	102.40%	
50) Toluene-d8	8.696	98	595741	49.79	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	99.58%	
62) 4-Bromofluorobenzene	11.122	95	208723	44.92	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	89.84%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.568	96	6916	1.91	ug/l	96
64) Tetrachloroethene	9.323	164	1265	0.36	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020321\
Data File : VX020851.D
Acq On : 03 Feb 2021 19:24
Operator : JC/SP
Sample : M1289-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-FD-02

Quant Time: Feb 04 02:27:00 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

