

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020322\
 Data File : VX026756.D
 Acq On : 03 Feb 2022 14:29
 Operator : JC/MD
 Sample : VX0203WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0203WBSD01

Quant Time: Feb 04 02:32:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	93491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	150956	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	131482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59548	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59512	45.568	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.140%
35) Dibromofluoromethane	5.391	113	48218	49.463	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.920%
50) Toluene-d8	8.647	98	175003	49.818	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.079	95	63176	50.626	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.260%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	15565	14.099	ug/l	98
3) Chloromethane	1.294	50	14672	14.498	ug/l	97
4) Vinyl Chloride	1.374	62	15402	14.495	ug/l	99
5) Bromomethane	1.611	94	7464	15.228	ug/l	98
6) Chloroethane	1.691	64	9706	18.064	ug/l	95
7) Trichlorofluoromethane	1.892	101	25987	15.491	ug/l	100
8) Diethyl Ether	2.136	74	9821	16.544	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.331	101	18629	18.222	ug/l	97
10) Methyl Iodide	2.459	142	20761	21.632	ug/l	97
11) Tert butyl alcohol	2.971	59	17119	94.227	ug/l	96
12) 1,1-Dichloroethene	2.325	96	16797	17.203	ug/l	90
13) Acrolein	2.239	56	16662	86.308	ug/l	98
14) Allyl chloride	2.666	41	30484	16.977	ug/l	96
15) Acrylonitrile	3.062	53	57614	103.018	ug/l	99
16) Acetone	2.386	43	54843	82.346	ug/l	97
17) Carbon Disulfide	2.514	76	38237	13.413	ug/l	97
18) Methyl Acetate	2.703	43	26609	20.097	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	64556	18.222	ug/l	99
20) Methylene Chloride	2.794	84	19919	18.102	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	17984	17.014	ug/l	97
22) Diisopropyl ether	3.763	45	67825	18.596	ug/l	95
23) Vinyl Acetate	3.721	43	282520	94.101	ug/l	100
24) 1,1-Dichloroethane	3.611	63	35649	18.098	ug/l	99
25) 2-Butanone	4.556	43	80122	92.506	ug/l	98
26) 2,2-Dichloropropane	4.477	77	25983	17.796	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	23078	19.130	ug/l	94
28) Bromochloromethane	4.903	49	16177	20.001	ug/l	98
29) Tetrahydrofuran	5.007	42	51845	100.786	ug/l	99
30) Chloroform	5.092	83	37319	18.234	ug/l	99
31) Cyclohexane	5.470	56	30870	16.025	ug/l	95
32) 1,1,1-Trichloroethane	5.391	97	31836	17.684	ug/l	100
36) 1,1-Dichloropropene	5.696	75	26337	17.114	ug/l	98
37) Ethyl Acetate	4.714	43	30769	19.765	ug/l	99
38) Carbon Tetrachloride	5.678	117	27855	17.597	ug/l	98
39) Methylcyclohexane	7.385	83	32298	17.400	ug/l	94
40) Benzene	6.037	78	78096	18.360	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020322\
 Data File : VX026756.D
 Acq On : 03 Feb 2022 14:29
 Operator : JC/MD
 Sample : VX0203WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0203WBSD01

Quant Time: Feb 04 02:32:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	17929	20.331	ug/l	98
42) 1,2-Dichloroethane	6.086	62	29604	17.789	ug/l	98
43) Isopropyl Acetate	6.342	43	49661	19.749	ug/l	99
44) Trichloroethene	7.129	130	23345	19.342	ug/l	96
45) 1,2-Dichloropropane	7.434	63	20858	19.645	ug/l	97
46) Dibromomethane	7.580	93	14252	19.042	ug/l	97
47) Bromodichloromethane	7.824	83	28321	18.727	ug/l	97
48) Methyl methacrylate	7.696	41	24146	19.376	ug/l	98
49) 1,4-Dioxane	7.677	88	7606	371.912	ug/l	96
51) 4-Methyl-2-Pentanone	8.574	43	155852	104.492	ug/l	98
52) Toluene	8.720	92	49437	18.953	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	28482	18.835	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	31423	19.346	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	21076	20.705	ug/l	97
56) Ethyl methacrylate	9.116	69	32262	20.682	ug/l	96
57) 1,3-Dichloropropane	9.311	76	35426	20.074	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.244	63	88504	108.199	ug/l	98
59) 2-Hexanone	9.433	43	115313	101.013	ug/l	98
60) Dibromochloromethane	9.525	129	21627	20.389	ug/l	98
61) 1,2-Dibromoethane	9.610	107	21435	20.209	ug/l	99
64) Tetrachloroethene	9.275	164	25263	20.340	ug/l	97
65) Chlorobenzene	10.079	112	53082	19.171	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	20124	19.561	ug/l	99
67) Ethyl Benzene	10.195	91	94249	18.731	ug/l	100
68) m/p-Xylenes	10.305	106	70689	37.541	ug/l	99
69) o-Xylene	10.640	106	34822	19.165	ug/l	98
70) Styrene	10.659	104	58289	19.660	ug/l	99
71) Bromoform	10.799	173	13788	18.970	ug/l #	99
73) Isopropylbenzene	10.963	105	92303	18.910	ug/l	98
74) N-amyl acetate	10.841	43	37075	19.675	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	25875	19.750	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	31037	23.672	ug/l	91
77) Bromobenzene	11.201	156	21682	19.390	ug/l	98
78) n-propylbenzene	11.305	91	105721	18.994	ug/l	99
79) 2-Chlorotoluene	11.366	91	62850	18.620	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	76970	19.127	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	7342	20.881	ug/l	98
82) 4-Chlorotoluene	11.457	91	72312	18.641	ug/l	97
83) tert-Butylbenzene	11.719	119	72751	19.366	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	77248	19.193	ug/l	98
85) sec-Butylbenzene	11.890	105	95673	19.690	ug/l	100
86) p-Isopropyltoluene	12.012	119	78910	19.357	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	40256	19.206	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	40980	19.316	ug/l	98
89) n-Butylbenzene	12.335	91	67476	19.288	ug/l	100
90) Hexachloroethane	12.542	117	11616	18.509	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	39664	20.025	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.945	75	5943	18.769	ug/l	96
93) 1,2,4-Trichlorobenzene	13.591	180	25265	19.895	ug/l	98
94) Hexachlorobutadiene	13.725	225	11024	20.564	ug/l	99
95) Naphthalene	13.780	128	87570	20.674	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	25459	20.385	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020322\
Data File : VX026756.D
Acq On : 03 Feb 2022 14:29
Operator : JC/MD
Sample : VX0203WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Quant Time: Feb 04 02:32:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:168 Resp: 93491
Ion Ratio Lower Upper
168 100
99 56.9 50.3 75.5



#2
Dichlorodifluoromethane
Concen: 14.099 ug/l
RT: 1.172 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 85 Resp: 15565
Ion Ratio Lower Upper
85 100
87 31.8 16.4 49.1









#3
Chloromethane
Concen: 14.498 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion: 50 Resp: 14672
Ion Ratio Lower Upper
50 100
52 34.4 26.1 39.1



#4
Vinyl Chloride
Concen: 14.495 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 62 Resp: 15402
Ion Ratio Lower Upper
62 100
64 33.3 26.2 39.2









#5
Bromomethane
Concen: 15.228 ug/l
RT: 1.611 min Scan# 80
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion: 94 Resp: 7464
Ion Ratio Lower Upper
94 100
96 93.1 75.9 113.9



#6
Chloroethane
Concen: 18.064 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 64 Resp: 9706
Ion Ratio Lower Upper
64 100
66 34.7 25.4 38.2









#7
 Trichlorofluoromethane
 Concen: 15.491 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX026756.D
 Acq: 03 Feb 2022 14:29




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0203WBSD01

Tgt Ion:101 Resp: 25987
 Ion Ratio Lower Upper
 101 100
 103 64.0 51.3 76.9



#8
 Diethyl Ether
 Concen: 16.544 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX026756.D
 Acq: 03 Feb 2022 14:29

Tgt Ion: 74 Resp: 9821
 Ion Ratio Lower Upper
 74 100
 45 100.0 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.222 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:101 Resp: 18629
Ion Ratio Lower Upper
101 100
85 42.3 35.3 52.9
151 79.0 61.4 92.2



#10
Methyl Iodide
Concen: 21.632 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion:142 Resp: 20761
Ion Ratio Lower Upper
142 100
127 45.5 38.1 57.1
141 14.4 11.2 16.8









#11
Tert butyl alcohol
Concen: 94.227 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.036 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01

Tgt Ion: 59 Resp: 17119
Ion Ratio Lower Upper
59 100
57 10.5 7.2 10.8



#12
1,1-Dichloroethene
Concen: 17.203 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29


Tgt Ion: 96 Resp: 16797
Ion Ratio Lower Upper
96 100
61 155.4 138.5 207.7
98 61.5 51.8 77.6








#13
Acrolein
Concen: 86.308 ug/l
RT: 2.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01







Tgt Ion: 56 Resp: 16662
Ion Ratio Lower Upper
56 100
55 69.8 56.9 85.3



#14
Allyl chloride
Concen: 16.977 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 41 Resp: 30484
Ion Ratio Lower Upper
41 100
39 63.3 54.2 81.4
76 32.6 25.2 37.8









#15
Acrylonitrile
Concen: 103.018 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion: 53 Resp: 57614
Ion Ratio Lower Upper
53 100
52 82.4 66.6 100.0
51 36.0 29.5 44.3



#16
Acetone
Concen: 82.346 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 43 Resp: 54843
Ion Ratio Lower Upper
43 100
58 30.5 23.0 34.6









#17
Carbon Disulfide
Concen: 13.413 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01

Tgt Ion: 76 Resp: 38237
Ion Ratio Lower Upper
76 100
78 8.5 7.5 11.3



#18
Methyl Acetate
Concen: 20.097 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 43 Resp: 26609
Ion Ratio Lower Upper
43 100
74 23.2 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 18.222 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01

Tgt Ion: 73 Resp: 64556
Ion Ratio Lower Upper
73 100
57 21.3 17.4 26.2



#20
Methylene Chloride
Concen: 18.102 ug/l
RT: 2.794 min Scan# 280
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 84 Resp: 19919
Ion Ratio Lower Upper
84 100
49 115.5 101.5 152.3
51 35.0 31.2 46.8
86 62.0 51.3 76.9









#21
trans-1,2-Dichloroethene
Concen: 17.014 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion: 96 Resp: 17984
Ion Ratio Lower Upper
96 100
61 139.6 112.9 169.3
98 66.5 49.5 74.3



#22
Diisopropyl ether
Concen: 18.596 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 45 Resp: 67825
Ion Ratio Lower Upper
45 100
43 70.5 60.6 91.0
87 26.3 22.0 33.0
59 10.8 9.0 13.4









#23
 Vinyl Acetate
 Concen: 94.101 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX026756.D
 Acq: 03 Feb 2022 14:29




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0203WBSD01


Tgt Ion: 43 Resp: 282520
 Ion Ratio Lower Upper
 43 100
 86 10.3 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 18.098 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.006 min
 Lab File: VX026756.D
 Acq: 03 Feb 2022 14:29

Tgt Ion: 63 Resp: 35649
 Ion Ratio Lower Upper
 63 100
 98 7.5 3.7 11.1
 100 4.4 2.4 7.2





#25

2-Butanone

Concen: 92.506 ug/l

RT: 4.556 min Scan# 5

Delta R.T. -0.006 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :





VX0203WBSD01

Tgt Ion: 43 Resp: 80122

Ion Ratio Lower Upper

43 100

72 24.4 20.4 30.6



#26

2,2-Dichloropropane

Concen: 17.796 ug/l

RT: 4.477 min Scan# 556

Delta R.T. -0.006 min

Lab File: VX026756.D




Acq: 03 Feb 2022 14:29


Tgt Ion: 77 Resp: 25983

Ion Ratio Lower Upper

77 100

97 23.2 11.0 33.0





#27

cis-1,2-Dichloroethene

Concen: 19.130 ug/l

RT: 4.489 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :

VX0203WBSD01





Tgt Ion: 96 Resp: 23078

Ion Ratio Lower Upper

96 100

61 138.6 0.0 296.6

98 62.6 0.0 128.8



#28

Bromochloromethane

Concen: 20.001 ug/l

RT: 4.903 min Scan# 626

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29




Tgt Ion: 49 Resp: 16177


Ion Ratio Lower Upper

49 100

129 1.0 0.0 4.2

130 80.0 62.4 93.6





#29

Tetrahydrofuran

Concen: 100.786 ug/l

RT: 5.007 min Scan# 64

Delta R.T. -0.006 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :

VX0203WBSD01





Tgt Ion: 42 Resp: 51845

Ion Ratio Lower Upper

42 100

72 44.0 34.6 52.0

71 40.6 31.8 47.8



#30

Chloroform

Concen: 18.234 ug/l

RT: 5.092 min Scan# 657

Delta R.T. -0.006 min

Lab File: VX026756.D




Acq: 03 Feb 2022 14:29


Tgt Ion: 83 Resp: 37319

Ion Ratio Lower Upper

83 100


85 63.4 51.6 77.4








#31
Cyclohexane
Concen: 16.025 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01






Tgt Ion: 56 Resp: 30870
Ion Ratio Lower Upper
56 100
69 31.4 25.8 38.6
84 80.5 68.9 103.3



#32
1,1,1-Trichloroethane
Concen: 17.684 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 97 Resp: 31836
Ion Ratio Lower Upper
97 100
99 64.4 51.3 76.9
61 43.4 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 45.568 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :





VX0203WBSD01

Tgt Ion: 65 Resp: 59512

Ion Ratio Lower Upper

65 100

67 51.2 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29




Tgt Ion: 114 Resp: 150956

Ion Ratio Lower Upper

114 100

63 20.0 0.0 42.8

88 17.0 0.0 33.0









#35
Dibromofluoromethane
Concen: 49.463 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:113 Resp: 48218
Ion Ratio Lower Upper
113 100
111 102.6 81.2 121.8
192 19.6 15.0 22.4



#36
1,1-Dichloropropene
Concen: 17.114 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 75 Resp: 26337
Ion Ratio Lower Upper
75 100
110 35.8 17.2 51.6
77 30.4 24.6 36.8





#37

Ethyl Acetate

Concen: 19.765 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.006 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :

VX0203WBSD01





Tgt Ion: 43 Resp: 30769

Ion Ratio Lower Upper

43 100

61 13.6 10.6 16.0

70 11.2 8.5 12.7



#38

Carbon Tetrachloride

Concen: 17.597 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29




Tgt Ion: 117 Resp: 27855

Ion Ratio Lower Upper

117 100

119 91.8 74.7 112.1

121 30.5 25.4 38.0





#39

Methylcyclohexane

Concen: 17.400 ug/l

RT: 7.385 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :

VX0203WBSD01





Tgt Ion: 83 Resp: 32298

Ion Ratio Lower Upper

83 100

55 73.7 63.6 95.4

98 44.7 38.4 57.6



#40

Benzene

Concen: 18.360 ug/l

RT: 6.037 min Scan# 812

Delta R.T. -0.006 min

Lab File: VX026756.D




Acq: 03 Feb 2022 14:29


Tgt Ion: 78 Resp: 78096

Ion Ratio Lower Upper

78 100

77 22.9 18.9 28.3









#41
Methacrylonitrile
Concen: 20.331 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:	41	Resp:	17929
Ion	Ratio	Lower	Upper
41	100		
39	54.4	45.4	68.0
67	66.9	53.2	79.8
52	27.6	23.0	34.6



#42
1,2-Dichloroethane
Concen: 17.789 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion:	62	Resp:	29604
Ion	Ratio	Lower	Upper
62	100		
98	9.6	0.0	17.8









#43
Isopropyl Acetate
Concen: 19.749 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01

Tgt Ion: 43 Resp: 49661
Ion Ratio Lower Upper
43 100
61 19.6 15.4 23.0
87 13.0 10.2 15.2



#44
Trichloroethene
Concen: 19.342 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion:130 Resp: 23345
Ion Ratio Lower Upper
130 100
95 97.6 0.0 202.6





#45

1,2-Dichloropropane

Concen: 19.645 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :





VX0203WBSD01

Tgt Ion: 63 Resp: 20858

Ion Ratio Lower Upper

63 100

65 31.6 23.8 35.6



#46

Dibromomethane

Concen: 19.042 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29




Tgt Ion: 93 Resp: 14252


Ion Ratio Lower Upper

93 100

95 83.4 66.1 99.1

174 103.1 78.7 118.1









#47
Bromodichloromethane
Concen: 18.727 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion: 83	Resp: 28321
Ion Ratio	Lower Upper
83 100	
85 67.1	51.4 77.0
127 9.0	6.9 10.3



#48
Methyl methacrylate
Concen: 19.376 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 41	Resp: 24146
Ion Ratio	Lower Upper
41 100	
69 79.5	62.2 93.4
39 55.3	45.4 68.0





#49

1,4-Dioxane

Concen: 371.912 ug/l

RT: 7.677 min Scan# 110

Delta R.T. -0.018 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :

VX0203WBSD01





Tgt Ion: 88 Resp: 7606

Ion Ratio Lower Upper

88 100

43 34.4 29.9 44.9

58 69.3 57.8 86.6



#50

Toluene-d8

Concen: 49.818 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX026756.D




Acq: 03 Feb 2022 14:29


Tgt Ion: 98 Resp: 175003

Ion Ratio Lower Upper

98 100

100 65.6 53.2 79.8









#51
4-Methyl-2-Pentanone
Concen: 104.492 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion: 43 Resp: 155852
Ion Ratio Lower Upper
43 100
58 39.4 30.5 45.7



#52
Toluene
Concen: 18.953 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 92 Resp: 49437
Ion Ratio Lower Upper
92 100
91 170.4 135.7 203.5









#53
t-1,3-Dichloropropene
Concen: 18.835 ug/l
RT: 8.976 min Scan# 1194
Delta R.T. -0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion: 75 Resp: 28482
Ion Ratio Lower Upper
75 100
77 31.5 25.7 38.5



#54
cis-1,3-Dichloropropene
Concen: 19.346 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 75 Resp: 31423
Ion Ratio Lower Upper
75 100
77 31.8 26.2 39.2
39 50.1 43.5 65.3





#55

1,1,2-Trichloroethane

Concen: 20.705 ug/l

RT: 9.153 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :

VX0203WBSD01

Tgt Ion: 97 Resp: 21076





Ion Ratio Lower Upper

97 100

83 82.9 68.4 102.6

85 53.7 44.0 66.0

99 65.0 49.8 74.6



#56

Ethyl methacrylate

Concen: 20.682 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29




Tgt Ion: 69 Resp: 32262


Ion Ratio Lower Upper

69 100

41 68.0 57.2 85.8

39 36.3 30.9 46.3





#57

1,3-Dichloropropane

Concen: 20.074 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :





VX0203WBSD01

Tgt Ion: 76 Resp: 35426

Ion Ratio Lower Upper

76 100

78 33.2 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 108.199 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX026756.D




Acq: 03 Feb 2022 14:29

Tgt Ion: 63 Resp: 88504

Ion Ratio Lower Upper

63 100

106 29.7 23.0 34.6





#59

2-Hexanone

Concen: 101.013 ug/l

RT: 9.433 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :





VX0203WBSD01

Tgt Ion: 43 Resp: 115313

Ion Ratio Lower Upper

43 100

58 54.4 26.4 79.2



#60

Dibromochloromethane

Concen: 20.389 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX026756.D




Acq: 03 Feb 2022 14:29

Tgt Ion:129 Resp: 21627

Ion Ratio Lower Upper

129 100

127 75.7 38.5 115.5









#61
1,2-Dibromoethane
Concen: 20.209 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:107 Resp: 21435
Ion Ratio Lower Upper
107 100
109 93.6 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 50.626 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 95 Resp: 63176
Ion Ratio Lower Upper
95 100
174 76.9 0.0 148.6
176 74.0 0.0 141.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:117 Resp: 131482
Ion Ratio Lower Upper
117 100
82 54.5 45.2 67.8
119 32.1 26.2 39.2



#64
Tetrachloroethene
Concen: 20.340 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion:164 Resp: 25263
Ion Ratio Lower Upper
164 100
166 125.4 101.8 152.6
129 99.3 83.3 124.9
131 95.3 78.0 117.0









#65
Chlorobenzene
Concen: 19.171 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:112 Resp: 53082
Ion Ratio Lower Upper
112 100
114 31.9 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 19.561 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion:131 Resp: 20124
Ion Ratio Lower Upper
131 100
133 95.0 47.9 143.8
119 64.4 31.7 95.1









#67
Ethyl Benzene
Concen: 18.731 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion: 91 Resp: 94249
Ion Ratio Lower Upper
91 100
106 30.8 24.7 37.1



#68
m/p-Xylenes
Concen: 37.541 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion:106 Resp: 70689
Ion Ratio Lower Upper
106 100
91 205.8 165.3 247.9









#69
o-Xylene
Concen: 19.165 ug/l
RT: 10.640 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:106 Resp: 34822
Ion Ratio Lower Upper
106 100
91 216.1 109.7 329.0



#70
Styrene
Concen: 19.660 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion:104 Resp: 58289
Ion Ratio Lower Upper
104 100
78 51.3 42.5 63.7
103 54.2 43.3 64.9









#71
Bromoform
Concen: 18.970 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:173 Resp: 13788
Ion Ratio Lower Upper
173 100
175 49.0 24.3 72.9
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion:152 Resp: 59548
Ion Ratio Lower Upper
152 100
115 70.6 44.2 132.6
150 164.0 0.0 348.2









#73
Isopropylbenzene
Concen: 18.910 ug/l
RT: 10.963 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:105 Resp: 92303
Ion Ratio Lower Upper
105 100
120 26.9 13.1 39.1



#74
N-ethyl acetate
Concen: 19.675 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 43 Resp: 37075
Ion Ratio Lower Upper
43 100
70 42.9 32.9 49.3
55 24.7 18.3 27.5
61 24.2 18.9 28.3





#75

1,1,2,2-Tetrachloroethane

Concen: 19.750 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :

VX0203WBSD01





Tgt Ion: 83 Resp: 25875

Ion Ratio Lower Upper

83 100

131 10.6 5.3 15.9

85 64.6 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 23.672 ug/l

RT: 11.244 min Scan# 1666

Delta R.T. -0.000 min

Lab File: VX026756.D




Acq: 03 Feb 2022 14:29


Tgt Ion: 75 Resp: 31037

Ion Ratio Lower Upper

75 100

77 31.2 18.3 54.8









#77
Bromobenzene
Concen: 19.390 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:156 Resp: 21682
Ion Ratio Lower Upper
156 100
77 165.1 84.5 253.4
158 98.2 48.1 144.4



#78
n-propylbenzene
Concen: 18.994 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 91 Resp: 105721
Ion Ratio Lower Upper
91 100
120 23.5 11.6 34.7









#79
2-Chlorotoluene
Concen: 18.620 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion: 91 Resp: 62850
Ion Ratio Lower Upper
91 100
126 34.3 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 19.127 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 105 Resp: 76970
Ion Ratio Lower Upper
105 100
120 49.0 24.0 72.0









#81
trans-1,4-Dichloro-2-butene
Concen: 20.881 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion: 75 Resp: 7342
Ion Ratio Lower Upper
75 100
53 117.8 94.0 141.0
89 42.8 37.0 55.4



#82
4-Chlorotoluene
Concen: 18.641 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 91 Resp: 72312
Ion Ratio Lower Upper
91 100
126 29.7 14.1 42.4





#83

tert-Butylbenzene

Concen: 19.366 ug/l

RT: 11.719 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :

VX0203WBSD01





Tgt Ion:119 Resp: 72751

Ion Ratio Lower Upper

119 100

91 61.8 31.4 94.0

134 23.8 11.8 35.3



#84

1,2,4-Trimethylbenzene

Concen: 19.193 ug/l

RT: 11.756 min Scan# 1750

Delta R.T. 0.006 min

Lab File: VX026756.D




Acq: 03 Feb 2022 14:29


Tgt Ion:105 Resp: 77248

Ion Ratio Lower Upper

105 100

120 44.8 21.9 65.5









#85
 sec-Butylbenzene
 Concen: 19.690 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX026756.D
 Acq: 03 Feb 2022 14:29




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0203WBSD01

Tgt Ion:105 Resp: 95673
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 19.357 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX026756.D
 Acq: 03 Feb 2022 14:29

Tgt Ion:119 Resp: 78910
 Ion Ratio Lower Upper
 119 100
 134 26.3 13.0 39.0
 91 24.5 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 19.206 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :

VX0203WBSD01





Tgt Ion:146 Resp: 40256

Ion Ratio Lower Upper

146 100

111 41.6 21.3 63.7

148 65.5 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 19.316 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29




Tgt Ion:146 Resp: 40980


Ion Ratio Lower Upper

146 100

111 40.2 20.2 60.6


148 64.5 31.4 94.0








#89
n-Butylbenzene
Concen: 19.288 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01



Tgt Ion: 91 Resp: 67476




Ion	Ratio	Lower	Upper
91	100		
92	54.1	27.2	81.6
134	26.2	13.1	39.1




#90
Hexachloroethane
Concen: 18.509 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 117 Resp: 11616

Ion	Ratio	Lower	Upper
117	100		
201	72.7	35.6	106.8









#91
1,2-Dichlorobenzene
Concen: 20.025 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29




Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01


Tgt Ion:146 Resp: 39664
Ion Ratio Lower Upper
146 100
111 43.1 21.8 65.3
148 64.3 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.769 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion: 75 Resp: 5943
Ion Ratio Lower Upper
75 100
155 83.2 40.0 119.9
157 108.2 51.6 154.8





#93

1,2,4-Trichlorobenzene

Concen: 19.895 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29

Instrument :

MSVOA_X

ClientSampleId :

VX0203WBSD01





Tgt Ion:180 Resp: 25265

Ion Ratio Lower Upper

180 100

182 95.4 46.7 140.0

145 33.3 16.3 48.8



#94

Hexachlorobutadiene

Concen: 20.564 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX026756.D

Acq: 03 Feb 2022 14:29




Tgt Ion:225 Resp: 11024


Ion Ratio Lower Upper

225 100

223 63.2 32.0 96.0

227 63.2 32.3 96.9









#95
Naphthalene
Concen: 20.674 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBSD01

Tgt Ion:128 Resp: 87570
Ion Ratio Lower Upper
128 100
127 12.6 10.3 15.5
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 20.385 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX026756.D
Acq: 03 Feb 2022 14:29

Tgt Ion:180 Resp: 25459
Ion Ratio Lower Upper
180 100
182 96.2 47.2 141.6
145 34.3 17.5 52.5

