

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020322\
 Data File : VX026758.D
 Acq On : 03 Feb 2022 15:16
 Operator : JC/MD
 Sample : PB142440ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142440ZHE#01

Quant Time: Feb 04 02:33:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	124141	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	207677	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179458	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79062	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78488	45.260	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	90.520%		
35) Dibromofluoromethane	5.391	113	65336	48.718	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.440%		
50) Toluene-d8	8.653	98	238230	49.294	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.580%		
62) 4-Bromofluorobenzene	11.085	95	84788	49.387	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.780%		
Target Compounds					Qvalue	
16) Acetone	2.386	43	8820	9.973	ug/l	89
20) Methylene Chloride	2.794	84	3245	2.221	ug/l	98
43) Isopropyl Acetate	6.348	43	22220	6.423	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020322\
Data File : VX026758.D
Acq On : 03 Feb 2022 15:16
Operator : JC/MD
Sample : PB142440ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142440ZHE#01

Quant Time: Feb 04 02:33:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

