

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020322\
 Data File : VX026760.D
 Acq On : 03 Feb 2022 16:02
 Operator : JC/MD
 Sample : PB142440ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142440ZHE#03

Quant Time: Feb 04 02:33:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	117058	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	201273	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75383	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	75931	46.435	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.860%
35) Dibromofluoromethane	5.391	113	62709	48.247	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.500%
50) Toluene-d8	8.653	98	226260	48.307	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.620%
62) 4-Bromofluorobenzene	11.085	95	81267	48.843	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.680%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	9021	10.818	ug/l	99
18) Methyl Acetate	2.715	43	2076	1.252	ug/l	98
20) Methylene Chloride	2.794	84	3265	2.370	ug/l	94
43) Isopropyl Acetate	6.348	43	22241	6.634	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020322\
Data File : VX026760.D
Acq On : 03 Feb 2022 16:02
Operator : JC/MD
Sample : PB142440ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142440ZHE#03

Quant Time: Feb 04 02:33:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

