

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
 Data File : VX044807.D
 Acq On : 03 Feb 2025 11:55
 Operator : JC/MD
 Sample : PB166357TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB166357TB

Quant Time: Feb 04 02:07:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	120503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	240248	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	214774	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	90113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89072	48.970	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.940%
35) Dibromofluoromethane	5.379	113	77998	51.104	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.200%
50) Toluene-d8	8.647	98	301141	56.587	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	113.180%#
62) 4-Bromofluorobenzene	11.079	95	103041	51.939	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.880%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	34521	47.970	ug/l	98
18) Methyl Acetate	2.703	43	2414	1.087	ug/l	91
20) Methylene Chloride	2.788	84	3544	2.115	ug/l #	94
43) Isopropyl Acetate	6.336	43	138341	33.238	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044807.D
Acq On : 03 Feb 2025 11:55
Operator : JC/MD
Sample : PB166357TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB166357TB

Quant Time: Feb 04 02:07:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

