

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
 Data File : VX044809.D
 Acq On : 03 Feb 2025 12:41
 Operator : JC/MD
 Sample : Q1216-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JPP-18.1-012825

Quant Time: Feb 04 02:08:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	123160	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	250036	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	218940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	93786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	90234	48.539	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.080%
35) Dibromofluoromethane	5.379	113	79443	50.013	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.020%
50) Toluene-d8	8.647	98	296369	53.510	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.020%
62) 4-Bromofluorobenzene	11.079	95	100988	48.911	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.820%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	35732	48.582	ug/l	94
18) Methyl Acetate	2.703	43	3903	1.720	ug/l	98
20) Methylene Chloride	2.782	84	2312	1.350	ug/l	97
43) Isopropyl Acetate	6.336	43	123523	28.516	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044809.D
Acq On : 03 Feb 2025 12:41
Operator : JC/MD
Sample : Q1216-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-18.1-012825

Quant Time: Feb 04 02:08:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

