

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
 Data File : VX044816.D
 Acq On : 03 Feb 2025 15:45
 Operator : JC/MD
 Sample : Q1232-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JPP-46.1-012925

Quant Time: Feb 04 02:10:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	117794	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	241488	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	213988	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	91814	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	83473	46.947	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.900%
35) Dibromofluoromethane	5.379	113	73061	47.624	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.240%
50) Toluene-d8	8.647	98	279762	52.300	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.600%
62) 4-Bromofluorobenzene	11.079	95	96576	48.430	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.860%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	36349	51.672	ug/l	99
18) Methyl Acetate	2.703	43	2645	1.219	ug/l #	83
20) Methylene Chloride	2.782	84	4134	2.524	ug/l	89
25) 2-Butanone	4.587	43	5253	5.022	ug/l	99
43) Isopropyl Acetate	6.336	43	130210	31.124	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044816.D
Acq On : 03 Feb 2025 15:45
Operator : JC/MD
Sample : Q1232-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-46.1-012925

Quant Time: Feb 04 02:10:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

