

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
 Data File : VX044818.D
 Acq On : 03 Feb 2025 16:32
 Operator : JC/MD
 Sample : Q1232-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JPP-42.2-012925

Quant Time: Feb 04 02:11:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	118473	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	237156	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	210118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	90092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82961	46.392	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.780%
35) Dibromofluoromethane	5.379	113	73382	48.706	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.420%
50) Toluene-d8	8.646	98	280393	53.375	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.740%
62) 4-Bromofluorobenzene	11.079	95	98683	50.391	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.780%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	36355	51.384	ug/l	97
18) Methyl Acetate	2.708	43	2564	1.175	ug/l #	86
20) Methylene Chloride	2.782	84	4091	2.483	ug/l #	92
43) Isopropyl Acetate	6.336	43	125480	30.541	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044818.D
Acq On : 03 Feb 2025 16:32
Operator : JC/MD
Sample : Q1232-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-42.2-012925

Quant Time: Feb 04 02:11:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

