

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020419\
 Data File : VX007353.D
 Acq On : 04 Feb 2019 12:30
 Operator : JC/SP
 Sample : K1296-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW202I-20190129

Quant Time: Feb 05 00:16:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	529538	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	821347	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	770275	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	372219	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	283692	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
35) Dibromofluoromethane	5.50	113	263170	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
50) Toluene-d8	8.71	98	980750	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
62) 4-Bromofluorobenzene	11.13	95	346868	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.93	101	22599	2.469	ug/l	100
12) 1,1-Dichloroethene	2.38	96	5180	1.025	ug/l	87
16) Acetone	2.45	43	7860	3.053	ug/l	92
19) Methyl tert-butyl Ether	3.20	73	56919	3.475	ug/l	98
29) Tetrahydrofuran	5.17	42	28874	12.530	ug/l	98
30) Chloroform	5.21	83	4514	0.497	ug/l #	71
44) Trichloroethene	7.21	130	6305	0.964	ug/l	99
64) Tetrachloroethene	9.34	164	10347	1.425	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020419\
Data File : VX007353.D
Acq On : 04 Feb 2019 12:30
Operator : JC/SP
Sample : K1296-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW202I-20190129

Quant Time: Feb 05 00:16:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

