

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020419\  
 Data File : VX007355.D  
 Acq On : 04 Feb 2019 13:23  
 Operator : JC/SP  
 Sample : PB116780TB  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB116780TB

Quant Time: Feb 05 00:21:24 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X020119W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Feb 02 01:32:36 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 499550   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.86  | 114  | 778085   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.11 | 117  | 757067   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.08 | 152  | 362525   | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.07  | 65   | 271498   | 51.20  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 102.40% |          |
| 35) Dibromofluoromethane    | 5.51  | 113  | 246541   | 48.70  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 97.40%  |          |
| 50) Toluene-d8              | 8.71  | 98   | 956843   | 50.94  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 101.88% |          |
| 62) 4-Bromofluorobenzene    | 11.13 | 95   | 337154   | 49.16  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 98.32%  |          |
| Target Compounds            |       |      |          |        |         |          |
| 16) Acetone                 | 2.45  | 43   | 34626    | 14.256 | ug/l    | 99       |
| 20) Methylene Chloride      | 2.86  | 84   | 53511    | 10.475 | ug/l    | 96       |
| 25) 2-Butanone              | 4.70  | 43   | 7590     | 2.196  | ug/l #  | 85       |
| 43) Isopropyl Acetate       | 6.46  | 43   | 32299    | 2.965  | ug/l    | 95       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020419\  
Data File : VX007355.D  
Acq On : 04 Feb 2019 13:23  
Operator : JC/SP  
Sample : PB116780TB  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
PB116780TB

Quant Time: Feb 05 00:21:24 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X020119W.M  
Quant Title : SW846 8260  
QLast Update : Sat Feb 02 01:32:36 2019  
Response via : Initial Calibration

