

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020419\
 Data File : VX007356.D
 Acq On : 04 Feb 2019 13:50
 Operator : JC/SP
 Sample : PB116780ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116780ZHE#01

Quant Time: Feb 05 00:23:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	495802	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	767333	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	761579	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	369998	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	267094	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
35) Dibromofluoromethane	5.50	113	244212	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	8.71	98	962095	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
62) 4-Bromofluorobenzene	11.13	95	346241	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
Target Compounds						
16) Acetone	2.45	43	35632	14.781	ug/l	94
18) Methyl Acetate	2.78	43	20824	3.869	ug/l	99
20) Methylene Chloride	2.86	84	32486	6.254	ug/l	99
25) 2-Butanone	4.70	43	9119	2.658	ug/l #	89
43) Isopropyl Acetate	6.46	43	34126	3.177	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020419\
Data File : VX007356.D
Acq On : 04 Feb 2019 13:50
Operator : JC/SP
Sample : PB116780ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116780ZHE#01

Quant Time: Feb 05 00:23:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

