

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020419\
 Data File : VX007358.D
 Acq On : 04 Feb 2019 14:43
 Operator : JC/SP
 Sample : PB116780ZHE#03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116780ZHE#03

Quant Time: Feb 05 00:38:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	486090	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	751560	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	727104	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	339885	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	259747	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
35) Dibromofluoromethane	5.50	113	233187	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
50) Toluene-d8	8.71	98	926429	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
62) 4-Bromofluorobenzene	11.13	95	323805	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
Target Compounds						
16) Acetone	2.45	43	34803	14.726	ug/l	96
18) Methyl Acetate	2.78	43	34523	6.133	ug/l	95
20) Methylene Chloride	2.85	84	67174	13.629	ug/l	96
25) 2-Butanone	4.71	43	8324	2.475	ug/l	98
43) Isopropyl Acetate	6.46	43	32146	3.055	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020419\
Data File : VX007358.D
Acq On : 04 Feb 2019 14:43
Operator : JC/SP
Sample : PB116780ZHE#03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116780ZHE#03

Quant Time: Feb 05 00:38:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

