

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020519\
 Data File : VX007376.D
 Acq On : 05 Feb 2019 14:58
 Operator : JC/SP
 Sample : PB116824ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 PB116824ZHE#02

Quant Time: Feb 06 01:07:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	481917	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	765616	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	732630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	355658	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	267992	52.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.78%	
35) Dibromofluoromethane	5.50	113	241163	48.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.82%	
50) Toluene-d8	8.71	98	942190	50.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.96%	
62) 4-Bromofluorobenzene	11.13	95	327415	48.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.04%	

Target Compounds

					Qvalue	
16) Acetone	2.45	43	26853	11.460	ug/l	91
18) Methyl Acetate	2.78	43	10782	2.338	ug/l #	78
20) Methylene Chloride	2.85	84	10597	1.835	ug/l #	85
25) 2-Butanone	4.71	43	11201	3.359	ug/l #	85
43) Isopropyl Acetate	6.46	43	26143	2.439	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020519\
 Data File : VX007376.D
 Acq On : 05 Feb 2019 14:58
 Operator : JC/SP
 Sample : PB116824ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116824ZHE#02

Quant Time: Feb 06 01:07:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

