

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020519\
 Data File : VX007378.D
 Acq On : 05 Feb 2019 15:51
 Operator : JC/SP
 Sample : PB116824ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116824ZHE#04

Quant Time: Feb 06 01:14:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	478443	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	736195	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	692659	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	347834	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	261964	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
35) Dibromofluoromethane	5.50	113	229186	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
50) Toluene-d8	8.71	98	906985	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
62) 4-Bromofluorobenzene	11.14	95	310463	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
Target Compounds						
16) Acetone	2.46	43	27055	11.630	ug/l	98
18) Methyl Acetate	2.79	43	34262	6.179	ug/l	98
20) Methylene Chloride	2.85	84	11611	2.067	ug/l #	89
25) 2-Butanone	4.71	43	12213	3.689	ug/l	98
43) Isopropyl Acetate	6.46	43	27361	2.655	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020519\
Data File : VX007378.D
Acq On : 05 Feb 2019 15:51
Operator : JC/SP
Sample : PB116824ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116824ZHE#04

Quant Time: Feb 06 01:14:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

