

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020519\
 Data File : VX007383.D
 Acq On : 05 Feb 2019 18:05
 Operator : JC/SP
 Sample : PB116824ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116824ZHE#01

Quant Time: Feb 06 01:34:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	441419	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	678853	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	652626	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	329613	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	242356	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
35) Dibromofluoromethane	5.50	113	211960	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
50) Toluene-d8	8.71	98	827607	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
62) 4-Bromofluorobenzene	11.14	95	300374	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
Target Compounds						
16) Acetone	2.48	43	26717	12.448	ug/l	97
18) Methyl Acetate	2.79	43	48436	9.152	ug/l	99
20) Methylene Chloride	2.85	84	10458	2.008	ug/l	94
25) 2-Butanone	4.73	43	11465	3.754	ug/l #	86
43) Isopropyl Acetate	6.46	43	26352	2.773	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020519\
Data File : VX007383.D
Acq On : 05 Feb 2019 18:05
Operator : JC/SP
Sample : PB116824ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116824ZHE#01

Quant Time: Feb 06 01:34:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

