

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020519\
 Data File : VX007366.D
 Acq On : 05 Feb 2019 10:32
 Operator : JC/SP
 Sample : VX0205WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0205WBL01

Quant Time: Feb 06 00:34:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	516965	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	800892	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	755231	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	367602	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	285509	52.03	ug/l	0.00
Spiked Amount						
			Recovery	=	104.06%	
35) Dibromofluoromethane	5.50	113	256557	49.23	ug/l	0.00
Spiked Amount						
			Recovery	=	98.46%	
50) Toluene-d8	8.71	98	959015	49.60	ug/l	0.00
Spiked Amount						
			Recovery	=	99.20%	
62) 4-Bromofluorobenzene	11.13	95	339950	48.16	ug/l	0.00
Spiked Amount						
			Recovery	=	96.32%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020519\
Data File : VX007366.D
Acq On : 05 Feb 2019 10:32
Operator : JC/SP
Sample : VX0205WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0205WBL01

Quant Time: Feb 06 00:34:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. 0.00 min

Lab File: VX007366.D

Acq: 05 Feb 2019 10:32

Instrument :

MSVOA_X

ClientSampleId :


VX0205WBL01

Tot Ion:168 Resp: 516965

Ion Ratio Lower Upper

168 100

99 50.2 37.9 56.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

200000



150000

100000

50000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 52.032 ug/l

RT: 6.06 min Scan# 805

Delta R.T. -0.01 min

Lab File: VX007366.D


Acq: 05 Feb 2019 10:32

Tgt Ion: 65 Resp: 285509

Ion Ratio Lower Upper

65 100

67 52.5 0.0 108.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

120000

100000

80000


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX007366.D

Acq: 05 Feb 2019 10:32

Instrument :

MSVOA_X

ClientSampleId :

VX0205WBL01


Tot Ion:114 Resp: 800892

Ion Ratio Lower Upper

114 100

63 18.2 0.0 35.6




88 14.1 0.0 29.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 49.233 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.01 min

Lab File: VX007366.D

Acq: 05 Feb 2019 10:32


Tgt Ion:113 Resp: 256557

Ion Ratio Lower Upper

113 100

111 102.1 81.8 122.8



192 22.7 18.5 27.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.602 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX007366.D

Acq: 05 Feb 2019 10:32

Instrument :

MSVOA_X

ClientSampleId :





VX0205WBL01

Tot Ion: 98 Resp: 959015

Ion Ratio Lower Upper

98 100

100 64.1 51.6 77.4



#62

4-Bromofluorobenzene

Concen: 48.159 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX007366.D

Acq: 05 Feb 2019 10:32




Tgt Ion: 95 Resp: 339950


Ion Ratio Lower Upper

95 100

174 93.5 0.0 189.2

176 90.6 0.0 186.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX007366.D
Acq: 05 Feb 2019 10:32

Instrument :
MSVOA_X
ClientSampleId :
VX0205WBL01

Tot Ion:117 Resp: 755231
Ion Ratio Lower Upper
117 100
82 52.9 42.1 63.1
119 31.3 26.5 39.7




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX007366.D
Acq: 05 Feb 2019 10:32

Tgt Ion:152 Resp: 367602
Ion Ratio Lower Upper
152 100
115 54.0 38.0 114.0
150 156.4 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

