

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
 Data File : VX020863.D
 Acq On : 05 Feb 2021 11:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 06 01:01:47 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	316007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	476843	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	458816	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	235446	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	181461	42.10	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.20%
35) Dibromofluoromethane	5.458	113	155720	47.68	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.36%
50) Toluene-d8	8.690	98	574246	47.96	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.92%
62) 4-Bromofluorobenzene	11.116	95	225963	48.59	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.18%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.185	85	222138	44.59	ug/l 97
3) Chloromethane	1.313	50	191798	43.15	ug/l 100
4) Vinyl Chloride	1.392	62	195056	44.01	ug/l 100
5) Bromomethane	1.624	94	94754	44.50	ug/l 100
6) Chloroethane	1.703	64	115847	46.18	ug/l 97
7) Trichlorofluoromethane	1.910	101	289350	48.13	ug/l 98
8) Diethyl Ether	2.166	74	101069	46.84	ug/l 98
9) 1,1,2-Trichlorotrifluo...	2.361	101	147492	48.06	ug/l 98
10) Methyl Iodide	2.489	142	184994	42.98	ug/l 92
11) Tert butyl alcohol	3.014	59	172812	214.77	ug/l 99
12) 1,1-Dichloroethene	2.355	96	140735	46.37	ug/l 98
13) Acrolein	2.270	56	158544	261.25	ug/l 98
14) Allyl chloride	2.703	41	241829	42.79	ug/l 96
15) Acrylonitrile	3.111	53	450592	221.24	ug/l 100
16) Acetone	2.416	43	391826	239.76	ug/l 99
17) Carbon Disulfide	2.544	76	407980	41.86	ug/l 99
18) Methyl Acetate	2.745	43	192991	46.22	ug/l 100
19) Methyl tert-butyl Ether	3.166	73	537777	47.37	ug/l 98
20) Methylene Chloride	2.831	84	181591	44.03	ug/l 98
21) trans-1,2-Dichloroethene	3.136	96	169085	44.60	ug/l 99
22) Diisopropyl ether	3.825	45	497855	44.82	ug/l 92
23) Vinyl Acetate	3.782	43	2070405	226.54	ug/l 100
24) 1,1-Dichloroethane	3.666	63	301603	44.92	ug/l 98
25) 2-Butanone	4.629	43	587345	229.46	ug/l 98
26) 2,2-Dichloropropane	4.544	77	261172	48.47	ug/l 99
27) cis-1,2-Dichloroethene	4.556	96	191040	45.95	ug/l 99
28) Bromochloromethane	4.977	49	115970	40.29	ug/l 96
29) Tetrahydrofuran	5.080	42	361564	219.32	ug/l 100
30) Chloroform	5.172	83	322709	46.75	ug/l 100
31) Cyclohexane	5.544	56	251831	43.45	ug/l 100
32) 1,1,1-Trichloroethane	5.452	97	289654	47.67	ug/l 99
36) 1,1-Dichloropropene	5.763	75	233969	48.89	ug/l 98
37) Ethyl Acetate	4.788	43	212836	46.91	ug/l 100
38) Carbon Tetrachloride	5.745	117	262914	52.25	ug/l 99
39) Methylcyclohexane	7.434	83	266665	48.89	ug/l 98
40) Benzene	6.105	78	693913	49.23	ug/l 99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
 Data File : VX020863.D
 Acq On : 05 Feb 2021 11:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 06 01:01:47 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.995	41	119838	47.19	ug/l	99
42) 1,2-Dichloroethane	6.153	62	245070	48.75	ug/l	100
43) Isopropyl Acetate	6.409	43	337080	46.60	ug/l	99
44) Trichloroethene	7.184	130	188247	50.19	ug/l	97
45) 1,2-Dichloropropane	7.482	63	178378	48.56	ug/l	98
46) Dibromomethane	7.635	93	123614	50.13	ug/l	99
47) Bromodichloromethane	7.873	83	263188	50.52	ug/l	100
48) Methyl methacrylate	7.745	41	178025	48.71	ug/l	97
49) 1,4-Dioxane	7.708	88	85727	1000.12	ug/l	99
51) 4-Methyl-2-Pentanone	8.616	43	1123183	246.05	ug/l	99
52) Toluene	8.763	92	451142	50.91	ug/l	99
53) t-1,3-Dichloropropene	9.019	75	276898	50.51	ug/l	100
54) cis-1,3-Dichloropropene	8.409	75	302003	51.09	ug/l	95
55) 1,1,2-Trichloroethane	9.189	97	189104	51.39	ug/l	99
56) Ethyl methacrylate	9.153	69	262904	44.94	ug/l	98
57) 1,3-Dichloropropane	9.348	76	309588	50.15	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.287	63	645248	253.42	ug/l	98
59) 2-Hexanone	9.470	43	869244	255.79	ug/l	100
60) Dibromochloromethane	9.561	129	214246	52.66	ug/l	99
61) 1,2-Dibromoethane	9.653	107	195511	51.34	ug/l	99
64) Tetrachloroethene	9.317	164	190014	54.02	ug/l	96
65) Chlorobenzene	10.116	112	489957	49.36	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.201	131	191713	52.03	ug/l	99
67) Ethyl Benzene	10.232	91	870380	50.57	ug/l	100
68) m/p-Xylenes	10.342	106	665819	103.74	ug/l	98
69) o-Xylene	10.683	106	313963	52.01	ug/l	99
70) Styrene	10.695	104	553579	53.41	ug/l	100
71) Bromoform	10.836	173	159651	53.97	ug/l	100
73) Isopropylbenzene	11.000	105	875020	51.42	ug/l	100
74) N-amyl acetate	10.878	43	298546	42.02	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.250	83	274387	47.86	ug/l	98
76) 1,2,3-Trichloropropane	11.274	75	332401	60.87	ug/l	87
77) Bromobenzene	11.238	156	220855	51.28	ug/l	95
78) n-propylbenzene	11.342	91	994789	50.63	ug/l	99
79) 2-Chlorotoluene	11.402	91	599590	49.96	ug/l	100
80) 1,3,5-Trimethylbenzene	11.488	105	752684	51.75	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.055	75	97384	49.97	ug/l	96
82) 4-Chlorotoluene	11.494	91	711433	50.09	ug/l	99
83) tert-Butylbenzene	11.756	119	717855	51.11	ug/l	98
84) 1,2,4-Trimethylbenzene	11.787	105	769739	52.47	ug/l	99
85) sec-Butylbenzene	11.927	105	847248	52.19	ug/l	99
86) p-Isopropyltoluene	12.049	119	799452	53.06	ug/l	99
87) 1,3-Dichlorobenzene	12.006	146	393577	50.15	ug/l	99
88) 1,4-Dichlorobenzene	12.079	146	394531	48.41	ug/l	99
89) n-Butylbenzene	12.372	91	670285	49.67	ug/l	99
90) Hexachloroethane	12.579	117	138166	50.93	ug/l	99
91) 1,2-Dichlorobenzene	12.372	146	379948	50.24	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.981	75	60185	47.36	ug/l	96
93) 1,2,4-Trichlorobenzene	13.628	180	248822	50.73	ug/l	97
94) Hexachlorobutadiene	13.762	225	109941	47.57	ug/l	99
95) Naphthalene	13.817	128	807679	45.02	ug/l	99
96) 1,2,3-Trichlorobenzene	14.000	180	251341	52.63	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
Data File : VX020863.D
Acq On : 05 Feb 2021 11:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Feb 06 01:01:47 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.623 min Scan# 744
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08


Instrument :
MSVOA_X
ClientSampled :





Tgt Ion:168 Resp: 316007
Ion Ratio Lower Upper
168 100
99 57.7 45.8 68.6




#2
Dichlorodifluoromethane
Concen: 44.59 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08



Tgt Ion: 85 Resp: 222138
Ion Ratio Lower Upper
85 100
87 31.0 16.4 49.2









#3
Chloromethane
Concen: 43.15 ug/l
RT: 1.313 min Scan# 37
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 50 Resp: 191798
Ion Ratio Lower Upper
50 100
52 33.0 26.4 39.6



#4
Vinyl Chloride
Concen: 44.01 ug/l
RT: 1.392 min Scan# 50
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 62 Resp: 195056
Ion Ratio Lower Upper
62 100
64 31.9 25.7 38.5









#5
Bromomethane
Concen: 44.50 ug/l
RT: 1.624 min Scan# 88
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 94 Resp: 94754
Ion Ratio Lower Upper
94 100
96 95.0 76.4 114.6



#6
Chloroethane
Concen: 46.18 ug/l
RT: 1.703 min Scan# 101
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 64 Resp: 115847
Ion Ratio Lower Upper
64 100
66 33.4 25.4 38.0









#7
Trichlorofluoromethane
Concen: 48.13 ug/l
RT: 1.910 min Scan# 135
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 289350
Ion Ratio Lower Upper
101 100
103 64.5 53.2 79.8



#8
Diethyl Ether
Concen: 46.84 ug/l
RT: 2.166 min Scan# 177
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 74 Resp: 101069
Ion Ratio Lower Upper
74 100
45 87.5 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.06 ug/l
RT: 2.361 min Scan# 209
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 147492
Ion Ratio Lower Upper
101 100
85 43.4 34.9 52.3
151 79.3 61.4 92.2



#10
Methyl Iodide
Concen: 42.98 ug/l
RT: 2.489 min Scan# 230
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08


Tgt Ion:142 Resp: 184994
Ion Ratio Lower Upper
142 100
127 48.6 33.3 49.9
141 14.5 11.4 17.0








#11
Tert butyl alcohol
Concen: 214.77 ug/l
RT: 3.014 min Scan# 316
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Instrument :
MSVOA_X
ClientSampled :







Tgt Ion: 59 Resp: 172812
Ion Ratio Lower Upper
59 100
57 10.0 7.8 11.8



#12
1,1-Dichloroethene
Concen: 46.37 ug/l
RT: 2.355 min Scan# 208
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 96 Resp: 140735
Ion Ratio Lower Upper
96 100
61 157.8 124.6 187.0
98 64.8 50.6 76.0









#13
Acrolein
Concen: 261.25 ug/l
RT: 2.270 min Scan# 194
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 56 Resp: 158544
Ion Ratio Lower Upper
56 100
55 68.9 56.4 84.6



#14
Allyl chloride
Concen: 42.79 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 41 Resp: 241829
Ion Ratio Lower Upper
41 100
39 59.4 51.6 77.4
76 38.0 30.5 45.7









#15
Acrylonitrile
Concen: 221.24 ug/l
RT: 3.111 min Scan# 332
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 53 Resp: 450592
Ion Ratio Lower Upper
53 100
52 81.8 65.3 97.9
51 35.6 28.2 42.2



#16
Acetone
Concen: 239.76 ug/l
RT: 2.416 min Scan# 218
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 43 Resp: 391826
Ion Ratio Lower Upper
43 100
58 34.2 27.8 41.8









#17
Carbon Disulfide
Concen: 41.86 ug/l
RT: 2.544 min Scan# 239
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 407980
Ion Ratio Lower Upper
76 100
78 8.7 7.2 10.8



#18
Methyl Acetate
Concen: 46.22 ug/l
RT: 2.745 min Scan# 272
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 43 Resp: 192991
Ion Ratio Lower Upper
43 100
74 27.6 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 47.37 ug/l
RT: 3.166 min Scan# 341
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 73 Resp: 537777
Ion Ratio Lower Upper
73 100
57 20.5 17.1 25.7



#20
Methylene Chloride
Concen: 44.03 ug/l
RT: 2.831 min Scan# 286
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08


Tgt Ion: 84 Resp: 181591
Ion Ratio Lower Upper
84 100
49 111.4 88.6 132.8
51 35.1 27.0 40.6
86 65.6 50.3 75.5








#21
trans-1,2-Dichloroethene
Concen: 44.60 ug/l
RT: 3.136 min Scan# 336
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Instrument :
MSVOA_X
ClientSampled :







Tgt Ion: 96 Resp: 169085
Ion Ratio Lower Upper
96 100
61 134.0 107.7 161.5
98 61.9 50.4 75.6



#22
Diisopropyl ether
Concen: 44.82 ug/l
RT: 3.825 min Scan# 449
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 45 Resp: 497855
Ion Ratio Lower Upper
45 100
43 62.3 58.0 87.0
87 31.5 25.4 38.0
59 12.8 10.2 15.2









#23
 Vinyl Acetate
 Concen: 226.54 ug/l
 RT: 3.782 min Scan# 442
 Delta R.T. -0.000 min
 Lab File: VX020863.D
 Acq: 05 Feb 2021 11:08




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 2070405
 Ion Ratio Lower Upper
 43 100
 86 12.4 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 44.92 ug/l
 RT: 3.666 min Scan# 423
 Delta R.T. 0.000 min
 Lab File: VX020863.D
 Acq: 05 Feb 2021 11:08

Tgt Ion: 63 Resp: 301603
 Ion Ratio Lower Upper
 63 100
 98 8.3 3.8 11.2
 100 4.8 2.5 7.5









#25
2-Butanone
Concen: 229.46 ug/l
RT: 4.629 min Scan# 581
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 587345
Ion Ratio Lower Upper
43 100
72 30.5 23.7 35.5



#26
2,2-Dichloropropane
Concen: 48.47 ug/l
RT: 4.544 min Scan# 567
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 77 Resp: 261172
Ion Ratio Lower Upper
77 100
97 22.9 11.7 35.1









#27
 cis-1,2-Dichloroethene
 Concen: 45.95 ug/l
 RT: 4.556 min Scan# 569
 Delta R.T. -0.006 min
 Lab File: VX020863.D
 Acq: 05 Feb 2021 11:08




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 96 Resp: 191040
 Ion Ratio Lower Upper
 96 100
 61 140.6 0.0 285.6
 98 64.9 0.0 129.6



#28
 Bromochloromethane
 Concen: 40.29 ug/l
 RT: 4.977 min Scan# 638
 Delta R.T. -0.000 min
 Lab File: VX020863.D
 Acq: 05 Feb 2021 11:08

Tgt Ion: 49 Resp: 115970
 Ion Ratio Lower Upper
 49 100
 129 2.2 0.0 4.0
 130 88.7 68.2 102.2





#29

Tetrahydrofuran

Concen: 219.32 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX020863.D

Acq: 05 Feb 2021 11:08





Tgt Ion: 42 Resp: 361564

Ion Ratio Lower Upper

42 100

72 52.2 41.7 62.5

71 47.8 38.4 57.6



#30

Chloroform

Concen: 46.75 ug/l

RT: 5.172 min Scan# 670

Delta R.T. -0.006 min

Lab File: VX020863.D




Acq: 05 Feb 2021 11:08


Tgt Ion: 83 Resp: 322709

Ion Ratio Lower Upper

83 100


85 64.8 52.0 78.0








#31
Cyclohexane
Concen: 43.45 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08


Instrument :
MSVOA_X
ClientSampled :





Tgt Ion: 56 Resp: 251831
Ion Ratio Lower Upper
56 100
69 31.6 25.3 37.9
84 88.9 71.4 107.0




#32
1,1,1-Trichloroethane
Concen: 47.67 ug/l
RT: 5.452 min Scan# 716
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08



Tgt Ion: 97 Resp: 289654
Ion Ratio Lower Upper
97 100
99 64.2 51.5 77.3
61 40.7 34.2 51.4





#33

1,2-Dichloroethane-d4

Concen: 42.10 ug/l

RT: 6.025 min Scan# 810

Delta R.T. 0.000 min

Lab File: VX020863.D





Acq: 05 Feb 2021 11:08

Tgt Ion: 65 Resp: 181461

Ion Ratio Lower Upper

65 100

67 52.7 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.824 min Scan# 941

Delta R.T. -0.006 min

Lab File: VX020863.D

Acq: 05 Feb 2021 11:08




Tgt Ion: 114 Resp: 476843


Ion Ratio Lower Upper

114 100

63 19.7 0.0 40.8

88 15.6 0.0 32.8









#35
Dibromofluoromethane
Concen: 47.68 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 155720
Ion Ratio Lower Upper
113 100
111 103.2 81.9 122.9
192 20.2 16.7 25.1



#36
1,1-Dichloropropene
Concen: 48.89 ug/l
RT: 5.763 min Scan# 767
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08


Tgt Ion: 75 Resp: 233969
Ion Ratio Lower Upper
75 100
110 35.1 18.1 54.4
77 30.4 25.3 37.9








#37
Ethyl Acetate
Concen: 46.91 ug/l
RT: 4.788 min Scan# 607
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 212836
Ion Ratio Lower Upper
43 100
61 15.8 12.8 19.2
70 13.3 10.5 15.7



#38
Carbon Tetrachloride
Concen: 52.25 ug/l
RT: 5.745 min Scan# 764
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08


Tgt Ion: 117 Resp: 262914
Ion Ratio Lower Upper
117 100
119 94.3 76.2 114.2
121 29.7 24.8 37.2








#39
Methylcyclohexane
Concen: 48.89 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Instrument :
MSVOA_X
ClientSampled :







Tgt Ion: 83 Resp: 266665
Ion Ratio Lower Upper
83 100
55 75.4 58.9 88.3
98 48.5 38.1 57.1



#40
Benzene
Concen: 49.23 ug/l
RT: 6.105 min Scan# 823
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 78 Resp: 693913
Ion Ratio Lower Upper
78 100
77 23.3 18.9 28.3









#41
Methacrylonitrile
Concen: 47.19 ug/l
RT: 4.995 min Scan# 641
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 41 Resp: 119838
Ion Ratio Lower Upper
41 100
39 53.6 42.5 63.7
67 82.5 65.0 97.6
52 32.2 25.3 37.9



#42
1,2-Dichloroethane
Concen: 48.75 ug/l
RT: 6.153 min Scan# 831
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 62 Resp: 245070
Ion Ratio Lower Upper
62 100
98 10.0 0.0 20.2









#43
Isopropyl Acetate
Concen: 46.60 ug/l
RT: 6.409 min Scan# 873
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 337080
Ion Ratio Lower Upper
43 100
61 23.1 18.2 27.4
87 15.9 12.5 18.7



#44
Trichloroethene
Concen: 50.19 ug/l
RT: 7.184 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion:130 Resp: 188247
Ion Ratio Lower Upper
130 100
95 95.8 0.0 198.0









#45
1,2-Dichloropropane
Concen: 48.56 ug/l
RT: 7.482 min Scan# 1049
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 63 Resp: 178378
Ion Ratio Lower Upper
63 100
65 30.8 25.7 38.5



#46
Dibromomethane
Concen: 50.13 ug/l
RT: 7.635 min Scan# 1074
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 93 Resp: 123614
Ion Ratio Lower Upper
93 100
95 82.1 66.8 100.2
174 105.0 83.1 124.7









#47
Bromodichloromethane
Concen: 50.52 ug/l
RT: 7.873 min Scan# 1113
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 263188
Ion Ratio Lower Upper
83 100
85 64.4 51.4 77.0
127 8.9 6.8 10.2



#48
Methyl methacrylate
Concen: 48.71 ug/l
RT: 7.745 min Scan# 1092
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 41 Resp: 178025
Ion Ratio Lower Upper
41 100
69 96.2 76.0 114.0
39 51.4 44.6 67.0









#49
1,4-Dioxane
Concen: 1000.12 ug/l
RT: 7.708 min Scan# 1086
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 88 Resp: 85727
Ion Ratio Lower Upper
88 100
43 28.5 23.2 34.8
58 70.0 55.0 82.4



#50
Toluene-d8
Concen: 47.96 ug/l
RT: 8.690 min Scan# 1247
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 98 Resp: 574246
Ion Ratio Lower Upper
98 100
100 65.5 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 246.05 ug/l
RT: 8.616 min Scan# 1235
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 1123183
Ion Ratio Lower Upper
43 100
58 44.8 35.4 53.0



#52
Toluene
Concen: 50.91 ug/l
RT: 8.763 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 92 Resp: 451142
Ion Ratio Lower Upper
92 100
91 170.4 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 50.51 ug/l
RT: 9.019 min Scan# 1301
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 276898
Ion Ratio Lower Upper
75 100
77 31.4 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 51.09 ug/l
RT: 8.409 min Scan# 1201
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 75 Resp: 302003
Ion Ratio Lower Upper
75 100
77 31.4 25.7 38.5
39 38.7 35.1 52.7









#55
1,1,2-Trichloroethane
Concen: 51.39 ug/l
RT: 9.189 min Scan# 1329
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 97	Resp: 189104
Ion Ratio	Lower Upper
97 100	
83 85.7	69.3 103.9
85 54.8	44.3 66.5
99 60.9	50.3 75.5



#56
Ethyl methacrylate
Concen: 44.94 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 69	Resp: 262904
Ion Ratio	Lower Upper
69 100	
41 58.1	46.8 70.2
39 28.0	24.7 37.1









#57

1,3-Dichloropropane
Concen: 50.15 ug/l
RT: 9.348 min Scan# 1355
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Instrument :
MSVOA_X
ClientSampled :




Tgt Ion: 76 Resp: 309588
Ion Ratio Lower Upper
76 100
78 32.6 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 253.42 ug/l
RT: 8.287 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 63 Resp: 645248
Ion Ratio Lower Upper
63 100
106 30.4 23.5 35.3









#59
2-Hexanone
Concen: 255.79 ug/l
RT: 9.470 min Scan# 1375
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 869244
Ion Ratio Lower Upper
43 100
58 62.2 30.9 92.8



#60
Dibromochloromethane
Concen: 52.66 ug/l
RT: 9.561 min Scan# 1390
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 129 Resp: 214246
Ion Ratio Lower Upper
129 100
127 76.3 38.4 115.2









#61
1,2-Dibromoethane
Concen: 51.34 ug/l
RT: 9.653 min Scan# 1405
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:107 Resp: 195511
Ion Ratio Lower Upper
107 100
109 94.1 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 48.59 ug/l
RT: 11.116 min Scan# 1645
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 95 Resp: 225963
Ion Ratio Lower Upper
95 100
174 80.1 0.0 154.6
176 75.5 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.092 min Scan# 1477
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 458816
Ion Ratio Lower Upper
117 100
82 55.7 47.4 71.2
119 32.3 26.6 39.8



#64
Tetrachloroethene
Concen: 54.02 ug/l
RT: 9.317 min Scan# 1350
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion:164 Resp: 190014
Ion Ratio Lower Upper
164 100
166 122.5 103.4 155.2
129 92.0 70.8 106.2
131 92.7 72.5 108.7









#65
Chlorobenzene
Concen: 49.36 ug/l
RT: 10.116 min Scan# 1481
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:112 Resp: 489957
Ion Ratio Lower Upper
112 100
114 32.3 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 52.03 ug/l
RT: 10.201 min Scan# 1495
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion:131 Resp: 191713
Ion Ratio Lower Upper
131 100
133 96.5 47.7 143.1
119 63.4 32.1 96.5









#67
Ethyl Benzene
Concen: 50.57 ug/l
RT: 10.232 min Scan# 1500
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 870380
Ion Ratio Lower Upper
91 100
106 30.8 24.6 37.0



#68
m/p-Xylenes
Concen: 103.74 ug/l
RT: 10.342 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion:106 Resp: 665819
Ion Ratio Lower Upper
106 100
91 206.7 162.8 244.2









#69
o-Xylene
Concen: 52.01 ug/l
RT: 10.683 min Scan# 1574
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:106 Resp: 313963
Ion Ratio Lower Upper
106 100
91 217.4 108.1 324.2



#70
Styrene
Concen: 53.41 ug/l
RT: 10.695 min Scan# 1576
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion:104 Resp: 553579
Ion Ratio Lower Upper
104 100
78 49.9 40.0 60.0
103 54.1 43.4 65.0





#71

Bromoform

Concen: 53.97 ug/l

RT: 10.836 min Scan# 1599





Delta R.T. 0.000 min

Lab File: VX020863.D

Acq: 05 Feb 2021 11:08

Tgt Ion:173 Resp: 159651

Ion	Ratio	Lower	Upper
173	100		
175	48.1	23.9	71.8
254	0.1	0.1	0.1



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.061 min Scan# 1800




Delta R.T. -0.000 min


Lab File: VX020863.D

Acq: 05 Feb 2021 11:08

Tgt Ion:152 Resp: 235446

Ion	Ratio	Lower	Upper
152	100		
115	87.0	41.1	123.3
150	176.2	0.0	349.0





#73

Isopropylbenzene

Concen: 51.42 ug/l

RT: 11.000 min Scan# 1626

Delta R.T. 0.000 min

Lab File: VX020863.D





Acq: 05 Feb 2021 11:08

Tgt Ion:105 Resp: 875020

Ion Ratio Lower Upper

105 100

120 26.5 13.2 39.6



#74

N-ethyl acetate

Concen: 42.02 ug/l

RT: 10.878 min Scan# 1606

Delta R.T. 0.000 min

Lab File: VX020863.D

Acq: 05 Feb 2021 11:08

Tgt Ion: 43 Resp: 298546




Ion Ratio Lower Upper


43 100

70 51.0 39.0 58.6

55 28.7 21.5 32.3

61 27.9 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 47.86 ug/l
RT: 11.250 min Scan# 1667
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 274387
Ion Ratio Lower Upper
83 100
131 10.5 5.1 15.2
85 65.0 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 60.87 ug/l
RT: 11.274 min Scan# 1671
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 75 Resp: 332401
Ion Ratio Lower Upper
75 100
77 30.6 19.1 57.5









#77
Bromobenzene
Concen: 51.28 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 220855
Ion Ratio Lower Upper
156 100
77 156.1 82.8 248.6
158 96.1 49.1 147.4



#78
n-propylbenzene
Concen: 50.63 ug/l
RT: 11.342 min Scan# 1682
Delta R.T. 0.001 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 91 Resp: 994789
Ion Ratio Lower Upper
91 100
120 23.4 11.5 34.5









#79
2-Chlorotoluene
Concen: 49.96 ug/l
RT: 11.402 min Scan# 1692
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 599590
Ion Ratio Lower Upper
91 100
126 33.3 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 51.75 ug/l
RT: 11.488 min Scan# 1706
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 105 Resp: 752684
Ion Ratio Lower Upper
105 100
120 48.2 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 49.97 ug/l
RT: 11.055 min Scan# 1635
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 97384
Ion Ratio Lower Upper
75 100
53 88.5 74.5 111.7
89 45.3 37.9 56.9



#82
4-Chlorotoluene
Concen: 50.09 ug/l
RT: 11.494 min Scan# 1707
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 91 Resp: 711433
Ion Ratio Lower Upper
91 100
126 29.0 14.9 44.5









#83
tert-Butylbenzene
Concen: 51.11 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:119 Resp: 717855
Ion Ratio Lower Upper
119 100
91 58.0 27.9 83.7
134 22.6 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 52.47 ug/l
RT: 11.787 min Scan# 1755
Delta R.T. -0.006 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion:105 Resp: 769739
Ion Ratio Lower Upper
105 100
120 44.2 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 52.19 ug/l
 RT: 11.927 min Scan# 1778
 Delta R.T. -0.000 min
 Lab File: VX020863.D
 Acq: 05 Feb 2021 11:08




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion:105 Resp: 847248
 Ion Ratio Lower Upper
 105 100
 134 19.5 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 53.06 ug/l
 RT: 12.049 min Scan# 1798
 Delta R.T. -0.000 min
 Lab File: VX020863.D
 Acq: 05 Feb 2021 11:08

Tgt Ion:119 Resp: 799452
 Ion Ratio Lower Upper
 119 100
 134 25.1 12.9 38.7
 91 23.6 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 50.15 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 393577
Ion Ratio Lower Upper
146 100
111 40.8 20.2 60.5
148 64.2 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 48.41 ug/l
RT: 12.079 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion:146 Resp: 394531
Ion Ratio Lower Upper
146 100
111 40.5 19.6 58.8
148 63.5 31.6 95.0









#89
n-Butylbenzene
Concen: 49.67 ug/l
RT: 12.372 min Scan# 1851
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 670285
Ion Ratio Lower Upper
91 100
92 54.1 27.4 82.0
134 26.5 13.2 39.5



#90
Hexachloroethane
Concen: 50.93 ug/l
RT: 12.579 min Scan# 1885
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 117 Resp: 138166
Ion Ratio Lower Upper
117 100
201 79.9 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 50.24 ug/l
RT: 12.372 min Scan# 1851
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 379948
Ion Ratio Lower Upper
146 100
111 43.6 20.8 62.2
148 64.9 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.36 ug/l
RT: 12.981 min Scan# 1951
Delta R.T. 0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion: 75 Resp: 60185
Ion Ratio Lower Upper
75 100
155 86.6 45.3 135.9
157 110.5 57.7 173.1









#93
 1,2,4-Trichlorobenzene
 Concen: 50.73 ug/l
 RT: 13.628 min Scan# 2057
 Delta R.T. -0.000 min
 Lab File: VX020863.D
 Acq: 05 Feb 2021 11:08




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion:180 Resp: 248822
 Ion Ratio Lower Upper
 180 100
 182 92.8 47.7 143.1
 145 31.4 15.0 45.0



#94
 Hexachlorobutadiene
 Concen: 47.57 ug/l
 RT: 13.762 min Scan# 2079
 Delta R.T. -0.000 min
 Lab File: VX020863.D
 Acq: 05 Feb 2021 11:08

Tgt Ion:225 Resp: 109941
 Ion Ratio Lower Upper
 225 100
 223 61.0 31.1 93.2
 227 62.1 31.6 95.0









#95
Naphthalene
Concen: 45.02 ug/l
RT: 13.817 min Scan# 2088
Delta R.T. -0.000 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:128 Resp: 807679
Ion Ratio Lower Upper
128 100
127 12.5 10.3 15.5
129 10.8 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 52.63 ug/l
RT: 14.000 min Scan# 2118
Delta R.T. 0.001 min
Lab File: VX020863.D
Acq: 05 Feb 2021 11:08

Tgt Ion:180 Resp: 251341
Ion Ratio Lower Upper
180 100
182 93.3 47.6 142.9
145 33.6 16.1 48.3

