

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
 Data File : VX020864.D
 Acq On : 05 Feb 2021 11:41
 Operator : JC/MD
 Sample : VX0205MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0205MBL01

Quant Time: Feb 06 01:02:43 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	284044	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	479866	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	465250	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	203071	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	189767	48.98	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.96%
35) Dibromofluoromethane	5.458	113	164024	49.90	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.80%
50) Toluene-d8	8.696	98	589524	48.92	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.84%
62) 4-Bromofluorobenzene	11.122	95	208115	44.47	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.94%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
Data File : VX020864.D
Acq On : 05 Feb 2021 11:41
Operator : JC/MD
Sample : VX0205MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0205MBL01

Quant Time: Feb 06 01:02:43 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

