

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
 Data File : VX020865.D
 Acq On : 05 Feb 2021 12:05
 Operator : JC/MD
 Sample : VX0205WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0205WBL01

Quant Time: Feb 06 01:03:07 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	274782	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	464310	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	444976	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	191209	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	185991	49.62	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.24%
35) Dibromofluoromethane	5.464	113	162436	51.07	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.14%
50) Toluene-d8	8.696	98	572354	49.09	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.18%
62) 4-Bromofluorobenzene	11.122	95	199136	43.98	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.96%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
Data File : VX020865.D
Acq On : 05 Feb 2021 12:05
Operator : JC/MD
Sample : VX0205WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0205WBL01

Quant Time: Feb 06 01:03:07 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

