

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
 Data File : VX020871.D
 Acq On : 05 Feb 2021 15:10
 Operator : JC/MD
 Sample : M1329-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Feb 06 01:51:18 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	281101	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	474598	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	462189	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	210726	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	187536	48.91	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.82%		
35) Dibromofluoromethane	5.464	113	162343	49.94	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.88%		
50) Toluene-d8	8.695	98	597435	50.13	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.26%		
62) 4-Bromofluorobenzene	11.122	95	212202	45.85	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.70%		
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.349	96	902	0.33	ug/l #	63
20) Methylene Chloride	2.837	84	1164	0.32	ug/l #	95
44) Trichloroethene	7.183	130	1851	0.50	ug/l	85
52) Toluene	8.762	92	2435	0.28	ug/l	99
53) t-1,3-Dichloropropene	9.025	75	2063	0.38	ug/l #	87
64) Tetrachloroethene	9.317	164	1366	0.39	ug/l #	84
65) Chlorobenzene	10.122	112	2411	0.24	ug/l #	87
68) m/p-Xylenes	10.341	106	2075	0.32	ug/l	100
70) Styrene	10.695	104	2794	0.27	ug/l #	87
80) 1,3,5-Trimethylbenzene	11.494	105	4957	0.38	ug/l	99
91) 1,2-Dichlorobenzene	12.378	146	3847	0.57	ug/l	97
93) 1,2,4-Trichlorobenzene	13.627	180	5049	1.15	ug/l	90
94) Hexachlorobutadiene	13.768	225	4548	2.20	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
Data File : VX020871.D
Acq On : 05 Feb 2021 15:10
Operator : JC/MD
Sample : M1329-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Feb 06 01:51:18 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

