

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
 Data File : VX020872.D
 Acq On : 05 Feb 2021 16:07
 Operator : JC/MD
 Sample : M1329-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Feb 06 01:06:35 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	298610	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	510613	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	502725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	211370	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	201011	49.35	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.70%
35) Dibromofluoromethane	5.458	113	177173	50.66	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.32%
50) Toluene-d8	8.690	98	635971	49.60	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.20%
62) 4-Bromofluorobenzene	11.122	95	225658	45.32	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.64%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020521\
Data File : VX020872.D
Acq On : 05 Feb 2021 16:07
Operator : JC/MD
Sample : M1329-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Feb 06 01:06:35 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

