

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020525\
 Data File : VX044841.D
 Acq On : 05 Feb 2025 14:17
 Operator : JC/MD
 Sample : Q1250-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-TB-20250127

Quant Time: Feb 06 04:38:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	103739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	213681	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	191975	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	81456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	77148	49.269	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.540%
35) Dibromofluoromethane	5.379	113	66065	48.667	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.340%
50) Toluene-d8	8.647	98	254438	53.755	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.520%
62) 4-Bromofluorobenzene	11.079	95	87720	49.713	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.420%
Target Compounds						
16) Acetone	2.386	43	732	1.182	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020525\
Data File : VX044841.D
Acq On : 05 Feb 2025 14:17
Operator : JC/MD
Sample : Q1250-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-TB-20250127

Quant Time: Feb 06 04:38:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

