

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020525\
 Data File : VX044846.D
 Acq On : 05 Feb 2025 16:13
 Operator : JC/MD
 Sample : Q1250-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-380-382

Quant Time: Feb 06 04:40:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	102126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	204882	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185437	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	80489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75149	48.750	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.500%		
35) Dibromofluoromethane	5.379	113	64194	49.320	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.640%		
50) Toluene-d8	8.646	98	242793	53.498	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	107.000%		
62) 4-Bromofluorobenzene	11.079	95	83786	49.523	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	99.040%		
Target Compounds						
16) Acetone	2.385	43	4069	6.672	ug/l #	87
17) Carbon Disulfide	2.501	76	2586	0.793	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020525\
Data File : VX044846.D
Acq On : 05 Feb 2025 16:13
Operator : JC/MD
Sample : Q1250-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-380-382

Quant Time: Feb 06 04:40:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

