

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020525\
 Data File : VX044849.D
 Acq On : 05 Feb 2025 17:22
 Operator : JC/MD
 Sample : Q1250-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-420-422

Quant Time: Feb 06 04:41:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	98319	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	198864	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	74477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	71921	48.463	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.920%
35) Dibromofluoromethane	5.385	113	61525	48.700	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.400%
50) Toluene-d8	8.647	98	232547	52.791	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.580%
62) 4-Bromofluorobenzene	11.079	95	79214	48.238	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.480%
Target Compounds						
16) Acetone	2.386	43	2649	4.512	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020525\
Data File : VX044849.D
Acq On : 05 Feb 2025 17:22
Operator : JC/MD
Sample : Q1250-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-420-422

Quant Time: Feb 06 04:41:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

