

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020525\
 Data File : VX044850.D
 Acq On : 05 Feb 2025 17:45
 Operator : JC/MD
 Sample : Q1250-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-300-302

Quant Time: Feb 06 04:42:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	101162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	208056	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	189576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	80127	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	76579	50.151	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.300%	
35) Dibromofluoromethane	5.379	113	65610	49.639	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.280%	
50) Toluene-d8	8.647	98	246823	53.556	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.120%	
62) 4-Bromofluorobenzene	11.079	95	86548	50.375	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 100.760%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3594	5.949	ug/l #	87
17) Carbon Disulfide	2.502	76	982	0.304	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020525\
Data File : VX044850.D
Acq On : 05 Feb 2025 17:45
Operator : JC/MD
Sample : Q1250-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-300-302

Quant Time: Feb 06 04:42:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

