

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020525\
 Data File : VX044835.D
 Acq On : 05 Feb 2025 10:41
 Operator : JC/MD
 Sample : VX0205WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0205WBL01

Quant Time: Feb 06 04:35:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	112270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	224960	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	207732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	90416	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	78734	46.461	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.920%
35) Dibromofluoromethane	5.379	113	69310	48.498	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.000%
50) Toluene-d8	8.641	98	263350	52.849	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.700%
62) 4-Bromofluorobenzene	11.079	95	94212	50.716	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.440%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020525\
Data File : VX044835.D
Acq On : 05 Feb 2025 10:41
Operator : JC/MD
Sample : VX0205WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0205WBL01

Quant Time: Feb 06 04:35:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX044835.D
Acq: 05 Feb 2025 10:41




Instrument :
MSVOA_X
ClientSampleId :
VX0205WBL01


Tgt Ion:168 Resp: 112270
Ion Ratio Lower Upper
168 100
99 62.4 51.1 76.7



#33
1,2-Dichloroethane-d4
Concen: 46.461 ug/l
RT: 5.946 min Scan# 797
Delta R.T. -0.006 min
Lab File: VX044835.D
Acq: 05 Feb 2025 10:41

Tgt Ion: 65 Resp: 78734
Ion Ratio Lower Upper
65 100
67 53.8 0.0 103.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. 0.000 min
Lab File: VX044835.D
Acq: 05 Feb 2025 10:41




Instrument :
MSVOA_X
ClientSampleId :
VX0205WBL01


Tgt Ion:114 Resp: 224960
Ion Ratio Lower Upper
114 100
63 20.7 0.0 43.0
88 15.7 0.0 35.4



#35
Dibromofluoromethane
Concen: 48.498 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX044835.D
Acq: 05 Feb 2025 10:41

Tgt Ion:113 Resp: 69310
Ion Ratio Lower Upper
113 100
111 101.6 81.6 122.4
192 17.8 14.4 21.6









#50
Toluene-d8
Concen: 52.849 ug/l
RT: 8.641 min Scan# 1239
Delta R.T. -0.006 min
Lab File: VX044835.D
Acq: 05 Feb 2025 10:41




Instrument :
MSVOA_X
ClientSampleId :
VX0205WBL01


Tgt Ion: 98 Resp: 263350
Ion Ratio Lower Upper
98 100
100 65.7 53.2 79.8



#62
4-Bromofluorobenzene
Concen: 50.716 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044835.D
Acq: 05 Feb 2025 10:41

Tgt Ion: 95 Resp: 94212
Ion Ratio Lower Upper
95 100
174 71.5 0.0 133.0
176 70.3 0.0 129.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX044835.D
Acq: 05 Feb 2025 10:41

Instrument :
MSVOA_X
ClientSampleId :
VX0205WBL01

Tgt Ion:117 Resp: 207732
Ion Ratio Lower Upper
117 100
82 55.3 47.6 71.4
119 31.5 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX044835.D
Acq: 05 Feb 2025 10:41

Tgt Ion:152 Resp: 90416
Ion Ratio Lower Upper
152 100
115 61.2 43.8 131.4
150 157.2 0.0 347.8

