

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020619\
 Data File : VX007395.D
 Acq On : 06 Feb 2019 13:18
 Operator : JC/SP
 Sample : K1380-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS027-190204

Quant Time: Feb 07 00:49:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	559255	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	837715	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	760289	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	387109	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	291242	49.06	ug/l	0.00
Spiked Amount			Recovery	=	98.12%	
35) Dibromofluoromethane	5.50	113	252096	46.25	ug/l	0.00
Spiked Amount			Recovery	=	92.50%	
50) Toluene-d8	8.71	98	996128	49.26	ug/l	0.00
Spiked Amount			Recovery	=	98.52%	
62) 4-Bromofluorobenzene	11.14	95	347019	47.00	ug/l	0.00
Spiked Amount			Recovery	=	94.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.48	43	14800	5.443	ug/l	97
27) cis-1,2-Dichloroethene	4.59	96	218345	35.144	ug/l	89
44) Trichloroethene	7.21	130	35051	5.256	ug/l	93
64) Tetrachloroethene	9.34	164	37533	5.238	ug/l	95
65) Chlorobenzene	10.14	112	9430	0.568	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020619\
Data File : VX007395.D
Acq On : 06 Feb 2019 13:18
Operator : JC/SP
Sample : K1380-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS027-190204

Quant Time: Feb 07 00:49:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

