

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020619\
 Data File : VX007397.D
 Acq On : 06 Feb 2019 14:11
 Operator : JC/SP
 Sample : K1386-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-190205

Quant Time: Feb 07 09:22:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	602027	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	900786	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	798529	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	392469	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	313587	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
35) Dibromofluoromethane	5.50	113	271958	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
50) Toluene-d8	8.71	98	1053559	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
62) 4-Bromofluorobenzene	11.14	95	357318	45.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.02%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	10763	1.609	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020619\
Data File : VX007397.D
Acq On : 06 Feb 2019 14:11
Operator : JC/SP
Sample : K1386-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-190205

Quant Time: Feb 07 09:22:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

