

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020619\
 Data File : VX007400.D
 Acq On : 06 Feb 2019 15:31
 Operator : JC/SP
 Sample : K1392-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Feb 07 01:28:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	578482	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	875341	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	827851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	402122	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	306208	49.87	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.74%	
35) Dibromofluoromethane	5.50	113	277531	48.73	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.46%	
50) Toluene-d8	8.71	98	1055574	49.95	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.90%	
62) 4-Bromofluorobenzene	11.13	95	374038	48.48	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.96%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020619\
 Data File : VX007400.D
 Acq On : 06 Feb 2019 15:31
 Operator : JC/SP
 Sample : K1392-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 Client Sampled :
 TRIP-BLANK

Quant Time: Feb 07 01:28:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

