

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020619\
 Data File : VX007401.D
 Acq On : 06 Feb 2019 15:57
 Operator : JC/SP
 Sample : K1380-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS036-190204

Quant Time: Feb 07 01:31:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	561270	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	856589	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	813281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	398103	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	300355	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
35) Dibromofluoromethane	5.51	113	264911	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
50) Toluene-d8	8.71	98	1031301	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
62) 4-Bromofluorobenzene	11.13	95	370536	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
Target Compounds						
16) Acetone	2.45	43	11528	4.224	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	44870	7.196	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020619\
Data File : VX007401.D
Acq On : 06 Feb 2019 15:57
Operator : JC/SP
Sample : K1380-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS036-190204

Quant Time: Feb 07 01:31:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

