

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020619\
 Data File : VX007403.D
 Acq On : 06 Feb 2019 16:51
 Operator : JC/SP
 Sample : K1392-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-9

Quant Time: Feb 07 01:35:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	543623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	844200	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	798115	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	389323	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	295216	51.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.32%	
35) Dibromofluoromethane	5.51	113	264636	48.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.36%	
50) Toluene-d8	8.71	98	1024143	50.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.50%	
62) 4-Bromofluorobenzene	11.13	95	358982	48.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.50%	
Target Compounds						
16) Acetone	2.45	43	14591	5.520	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020619\
Data File : VX007403.D
Acq On : 06 Feb 2019 16:51
Operator : JC/SP
Sample : K1392-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-9

Quant Time: Feb 07 01:35:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

