

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020619\
 Data File : VX007408.D
 Acq On : 06 Feb 2019 19:04
 Operator : JC/SP
 Sample : K1395-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Feb 07 09:46:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	559961	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	869122	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	800770	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	385787	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	303996	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
35) Dibromofluoromethane	5.51	113	273644	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
50) Toluene-d8	8.71	98	1021533	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
62) 4-Bromofluorobenzene	11.13	95	362469	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	15397	2.475	ug/l	86
44) Trichloroethene	7.21	130	4745	0.686	ug/l	88
64) Tetrachloroethene	9.34	164	349489	46.305	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020619\
Data File : VX007408.D
Acq On : 06 Feb 2019 19:04
Operator : JC/SP
Sample : K1395-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Feb 07 09:46:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

