

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020619\
 Data File : VX007409.D
 Acq On : 06 Feb 2019 19:30
 Operator : JC/SP
 Sample : K1395-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12

Quant Time: Feb 07 02:33:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	541221	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	838030	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	784371	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	381315	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	290754	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
35) Dibromofluoromethane	5.51	113	263258	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
50) Toluene-d8	8.71	98	1000684	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
62) 4-Bromofluorobenzene	11.13	95	355125	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
Target Compounds						
16) Acetone	2.45	43	13696	5.205	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	8534	1.419	ug/l	90
64) Tetrachloroethene	9.34	164	183691	24.847	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020619\
Data File : VX007409.D
Acq On : 06 Feb 2019 19:30
Operator : JC/SP
Sample : K1395-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12

Quant Time: Feb 07 02:33:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

