

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX020619\
 Data File : VX007410.D
 Acq On : 06 Feb 2019 19:57
 Operator : JC/SP
 Sample : K1392-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Manual Integrations
 APPROVED

MMDadoda
 2/7/2019 9:53:11 AM

Quant Time: Feb 07 09:43:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	536250	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	823446	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	791266	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	406638	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	289043	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
35) Dibromofluoromethane	5.51	113	260120	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
50) Toluene-d8	8.71	98	997805	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	11.13	95	372031	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	54334	9.169	ug/l	97
16) Acetone	2.45	43	21962	8.423	ug/l	95
17) Carbon Disulfide	2.57	76	7192	2.635	ug/l	96
21) trans-1,2-Dichloroethene	3.17	96	16080	2.948	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	1562667	262.310	ug/l	91
31) Cyclohexane	5.58	56	7831	0.989	ug/l #	72
39) Methylcyclohexane	7.46	83	12252	1.371	ug/l	91
40) Benzene	6.14	78	32414	1.489	ug/l #	85
44) Trichloroethene	7.21	130	3463	0.528	ug/l	80
52) Toluene	8.79	92	619743	44.136	ug/l	98
64) Tetrachloroethene	9.34	164	3724	0.499	ug/l	91
67) Ethyl Benzene	10.25	91	241506	8.685	ug/l	98
68) m/p-Xylenes	10.36	106	668594	60.488	ug/l	99
69) o-Xylene	10.69	106	415321	39.048	ug/l	100
73) Isopropylbenzene	11.02	105	27954	1.012	ug/l	99
78) n-propylbenzene	11.36	91	75205	2.482	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	699058	30.922	ug/l	99
83) tert-Butylbenzene	11.77	119	43272	1.933	ug/l	95
84) 1,2,4-Trimethylbenzene	11.80	105	1052126	45.841	ug/l	98
85) sec-Butylbenzene	11.94	105	26669m	0.984	ug/l	
86) p-Isopropyltoluene	12.06	119	80777	3.303	ug/l	98
95) Naphthalene	13.83	128	353329	12.884	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020619\
Data File : VX007410.D
Acq On : 06 Feb 2019 19:57
Operator : JC/SP
Sample : K1392-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Manual Integrations
APPROVED

MMDadoda
2/7/2019 9:53:11 AM

Quant Time: Feb 07 09:43:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

