

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020619\
 Data File : VX007391.D
 Acq On : 06 Feb 2019 11:21
 Operator : JC/SP
 Sample : VX0206WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0206WBL01

Quant Time: Feb 07 00:37:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	625872	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	935330	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	833564	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	366099	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	330635	49.77	ug/l	0.00
Spiked Amount			Recovery	=	99.54%	
35) Dibromofluoromethane	5.50	113	298472	49.04	ug/l	0.00
Spiked Amount			Recovery	=	98.08%	
50) Toluene-d8	8.71	98	1083357	47.98	ug/l	0.00
Spiked Amount			Recovery	=	95.96%	
62) 4-Bromofluorobenzene	11.13	95	357893	43.41	ug/l	0.00
Spiked Amount			Recovery	=	86.82%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020619\
Data File : VX007391.D
Acq On : 06 Feb 2019 11:21
Operator : JC/SP
Sample : VX0206WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0206WBL01

Quant Time: Feb 07 00:37:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.00 min

Lab File: VX007391.D

Acq: 06 Feb 2019 11:21

Instrument :

MSVOA_X

ClientSampleId :


VX0206WBL01

Tot Ion:168 Resp: 625872

Ion Ratio Lower Upper




168 100

99 48.4 37.9 56.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 49.771 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.00 min

Lab File: VX007391.D


Acq: 06 Feb 2019 11:21

Tgt Ion: 65 Resp: 330635

Ion Ratio Lower Upper



65 100


67 54.4 0.0 108.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX007391.D

Acq: 06 Feb 2019 11:21

Instrument :

MSVOA_X

ClientSampleId :

VX0206WBL01


Tot Ion:114 Resp: 935330

Ion Ratio Lower Upper

114 100

63 17.8 0.0 35.6




88 14.7 0.0 29.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 49.044 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.01 min

Lab File: VX007391.D

Acq: 06 Feb 2019 11:21


Tgt Ion:113 Resp: 298472

Ion Ratio Lower Upper

113 100

111 101.6 81.8 122.8



192 23.5 18.5 27.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 47.979 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX007391.D
Acq: 06 Feb 2019 11:21




Instrument :
MSVOA_X
ClientSampleId :
VX0206WBL01


Tot Ion: 98 Resp: 1083357
Ion Ratio Lower Upper
98 100
100 64.5 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 43.413 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.00 min
Lab File: VX007391.D
Acq: 06 Feb 2019 11:21

Tgt Ion: 95 Resp: 357893
Ion Ratio Lower Upper
95 100
174 96.2 0.0 189.2
176 93.4 0.0 186.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX007391.D
Acq: 06 Feb 2019 11:21

Instrument :
MSVOA_X
ClientSampleId :
VX0206WBL01

Tot Ion:117 Resp: 833564
Ion Ratio Lower Upper
117 100
82 52.4 42.1 63.1
119 32.1 26.5 39.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX007391.D
Acq: 06 Feb 2019 11:21

Tgt Ion:152 Resp: 366099
Ion Ratio Lower Upper
152 100
115 54.4 38.0 114.0
150 156.1 0.0 346.4

