

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX020620\
 Data File : VX014846.D
 Acq On : 06 Feb 2020 14:37
 Operator : JC/SP
 Sample : L1346-05MEDL 5X
 Misc : 6.46g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PD-3-(35-37)MEDL

Quant Time: Feb 07 01:07:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	439625	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	674634	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	642046	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	335085	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	243182	47.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.74%	
35) Dibromofluoromethane	5.49	113	196352	46.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.26%	
50) Toluene-d8	8.71	98	814496	49.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.54%	
62) 4-Bromofluorobenzene	11.14	95	312332	53.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.16%	
Target Compounds						
31) Cyclohexane	5.60	56	83350	10.441	ug/l #	44
39) Methylcyclohexane	7.46	83	213959	27.245	ug/l	97
73) Isopropylbenzene	11.01	105	49361	2.095	ug/l	96
78) n-propylbenzene	11.35	91	299724	11.207	ug/l	98
85) sec-Butylbenzene	11.94	105	43584	1.936	ug/l	98
89) n-Butylbenzene	12.38	91	105621	5.829	ug/l #	76
95) Naphthalene	13.83	128	51254	3.073	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020620\
Data File : VX014846.D
Acq On : 06 Feb 2020 14:37
Operator : JC/SP
Sample : L1346-05MEDL 5X
Misc : 6.46g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PD-3-(35-37)MEDL

Quant Time: Feb 07 01:07:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

