

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020620\
 Data File : VX014850.D
 Acq On : 06 Feb 2020 16:10
 Operator : JC/SP
 Sample : L1430-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 N29305

Quant Time: Feb 07 02:18:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	455631	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	698463	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	646479	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	315951	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	252867	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
35) Dibromofluoromethane	5.49	113	207890	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
50) Toluene-d8	8.71	98	823489	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	11.13	95	295108	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
Target Compounds						
16) Acetone	2.43	43	17134	6.022	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020620\
Data File : VX014850.D
Acq On : 06 Feb 2020 16:10
Operator : JC/SP
Sample : L1430-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
N29305

Quant Time: Feb 07 02:18:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

