

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020620\
 Data File : VX014852.D
 Acq On : 06 Feb 2020 16:57
 Operator : JC/SP
 Sample : L1419-02 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30576

Quant Time: Feb 07 02:38:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	464991	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	716611	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	668284	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	320881	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	264479	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
35) Dibromofluoromethane	5.49	113	215913	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
50) Toluene-d8	8.71	98	853547	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
62) 4-Bromofluorobenzene	11.13	95	301872	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
Target Compounds						
16) Acetone	2.43	43	396423	133.277	ug/l	98
52) Toluene	8.78	92	15001	1.186	ug/l	99
68) m/p-Xylenes	10.35	106	3867	0.418	ug/l	82
95) Naphthalene	13.83	128	27754	2.142	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020620\
Data File : VX014852.D
Acq On : 06 Feb 2020 16:57
Operator : JC/SP
Sample : L1419-02 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30576

Quant Time: Feb 07 02:38:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

