

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020624\
 Data File : VX040136.D
 Acq On : 06 Feb 2024 14:07
 Operator : JC/MD
 Sample : P1376-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-700-702

Quant Time: Feb 07 00:34:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	171939	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158222	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80115	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63742	45.616	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.240%
35) Dibromofluoromethane	5.385	113	49398	42.204	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.400%
50) Toluene-d8	8.647	98	201549	46.046	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.100%
62) 4-Bromofluorobenzene	11.079	95	81745	48.501	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.000%
Target Compounds						
16) Acetone	2.374	43	2383	3.640	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020624\
Data File : VX040136.D
Acq On : 06 Feb 2024 14:07
Operator : JC/MD
Sample : P1376-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-700-702

Quant Time: Feb 07 00:34:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

