

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020624\
 Data File : VX040137.D
 Acq On : 06 Feb 2024 14:30
 Operator : JC/MD
 Sample : P1376-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-725-727

Quant Time: Feb 07 00:35:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	83296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	166660	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	154326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66756	49.798	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.600%		
35) Dibromofluoromethane	5.385	113	48123	42.417	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	84.840%		
50) Toluene-d8	8.647	98	200289	47.207	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.420%		
62) 4-Bromofluorobenzene	11.079	95	77739	47.585	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.180%		
Target Compounds						
16) Acetone	2.380	43	1503	2.393	ug/l	92
17) Carbon Disulfide	2.508	76	1472	0.537	ug/l #	91
94) Hexachlorobutadiene	13.719	225	283	0.468	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020624\
Data File : VX040137.D
Acq On : 06 Feb 2024 14:30
Operator : JC/MD
Sample : P1376-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-725-727

Quant Time: Feb 07 00:35:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

