

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020624\
 Data File : VX040141.D
 Acq On : 06 Feb 2024 16:01
 Operator : JC/MD
 Sample : P1376-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-740-742

Quant Time: Feb 07 00:37:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	56812	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	150723	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80055	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51021	55.802	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.600%
35) Dibromofluoromethane	5.391	113	37127	36.185	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	72.380%#
50) Toluene-d8	8.647	98	142046	37.020	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	74.040%#
62) 4-Bromofluorobenzene	11.079	95	79987	54.138	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.280%
Target Compounds						
16) Acetone	2.379	43	1865	4.353	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020624\
Data File : VX040141.D
Acq On : 06 Feb 2024 16:01
Operator : JC/MD
Sample : P1376-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-740-742

Quant Time: Feb 07 00:37:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

