

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020624\
 Data File : VX040143.D
 Acq On : 06 Feb 2024 16:46
 Operator : JC/MD
 Sample : P1376-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-DUP-20240202RE

Quant Time: Feb 07 01:26:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	78590	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	111798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	142280	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57970	45.833	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.660%
35) Dibromofluoromethane	5.379	113	41268	54.225	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.460%
50) Toluene-d8	8.647	98	133842	47.026	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.060%
62) 4-Bromofluorobenzene	11.079	95	73988	67.513	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	135.020%#
Target Compounds						
16) Acetone	2.380	43	1416	2.389	ug/l	Qvalue # 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020624\
Data File : VX040143.D
Acq On : 06 Feb 2024 16:46
Operator : JC/MD
Sample : P1376-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-DUP-20240202RE

Quant Time: Feb 07 01:26:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

