



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020624\
Data File : VX040130.D
Acq On : 06 Feb 2024 09:52
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Title : SW846 8260
Last Update : Thu Feb 01 05:44:39 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	7235	PASS
75	95	30	60	52.5	19392	PASS
95	95	100	100	100.0	36939	PASS
96	95	5	9	7.3	2698	PASS
173	174	0.00	2	0.9	238	PASS
174	95	50	100	70.9	26179	PASS
175	174	5	9	7.5	1955	PASS
176	174	95	101	97.1	25432	PASS
177	176	5	9	6.9	1767	PASS