

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020719\
 Data File : VX007425.D
 Acq On : 07 Feb 2019 15:09
 Operator : JC/SP
 Sample : K1401-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR-200050

Quant Time: Feb 08 22:32:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	522735	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	814327	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	794583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	386069	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	285133	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
35) Dibromofluoromethane	5.51	113	260158	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
50) Toluene-d8	8.71	98	1000193	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
62) 4-Bromofluorobenzene	11.13	95	360432	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
Target Compounds						
16) Acetone	2.45	43	42714	16.806	ug/l	100
18) Methyl Acetate	2.78	43	13525	2.611	ug/l	92
20) Methylene Chloride	2.86	84	626663	121.273	ug/l	97
25) 2-Butanone	4.71	43	13127	3.629	ug/l #	80
43) Isopropyl Acetate	6.46	43	20956	1.838	ug/l #	90
95) Naphthalene	13.83	128	222733	8.555	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020719\
Data File : VX007425.D
Acq On : 07 Feb 2019 15:09
Operator : JC/SP
Sample : K1401-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR-200050

Quant Time: Feb 08 22:32:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

