

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020719\
 Data File : VX007430.D
 Acq On : 07 Feb 2019 17:22
 Operator : JC/SP
 Sample : K1385-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 98-WC-S-01-2-57

Quant Time: Feb 08 22:41:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	497722	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	778216	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	762851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	373906	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	268646	50.85	ug/l	0.00
Spiked Amount			Recovery	=	101.70%	
35) Dibromofluoromethane	5.51	113	248471	49.07	ug/l	0.00
Spiked Amount			Recovery	=	98.14%	
50) Toluene-d8	8.71	98	958642	51.03	ug/l	0.00
Spiked Amount			Recovery	=	102.06%	
62) 4-Bromofluorobenzene	11.13	95	347578	50.67	ug/l	0.00
Spiked Amount			Recovery	=	101.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	21030	8.690	ug/l	99
18) Methyl Acetate	2.78	43	24918	4.498	ug/l	99
20) Methylene Chloride	2.86	84	5510177	1123.194	ug/l	97
25) 2-Butanone	4.71	43	7591	2.204	ug/l #	87
43) Isopropyl Acetate	6.46	43	27473	2.522	ug/l	97
95) Naphthalene	13.83	128	49890	1.978	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020719\
Data File : VX007430.D
Acq On : 07 Feb 2019 17:22
Operator : JC/SP
Sample : K1385-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
98-WC-S-01-2-57

Quant Time: Feb 08 22:41:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

