

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020720\
 Data File : VX014868.D
 Acq On : 07 Feb 2020 14:04
 Operator : JC/SP
 Sample : L1420-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 800-TRIP-BLANK

Quant Time: Feb 07 22:39:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jan 15 12:41:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	69176	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	383429	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	341994	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.06	65	153324	31.53	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	105.10%
60) 4-Bromofluorobenzene	11.13	95	156945	28.68	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	95.60%
63) Toluene-d8	8.71	98	452236	29.28	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	97.60%

Target Compounds					Ovalue
15) Acetone	2.43	58	16657	20.990	ug/l 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020720\
Data File : VX014868.D
Acq On : 07 Feb 2020 14:04
Operator : JC/SP
Sample : L1420-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
800-TRIP-BLANK

Quant Time: Feb 07 22:39:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jan 15 12:41:12 2020
Response via : Initial Calibration

